



AGENCE DES NORMES ET DE LA QUALITÉ®
STANDARDS AND QUALITY AGENCY®

PROJET DE NORME CAMEROUNAISE PNC 6756 : 2026, ISO 4214 :2022 - CT 48

2026

www.anor.cm

NORMES POUR LE LAIT ET PRODUITS LAITIERS — DETERMINATION DES ACIDES AMINES DANS LES FORMULES INFANTILES, LES PRODUITS NUTRITIONNELS POUR ADULTES/ENFANTS ET AUTRES PRODUITS LAITIERS

Toute reproduction ou représentation intégrale ou partielle, par quelque procédé que ce soit, des pages publiées dans le présent document, faite sans l'autorisation de l'éditeur est illicite et constitue une contrefaçon. Seules sont autorisées d'une part, les reproductions strictement réservées à l'usage privé du copiste et non destinées à une utilisation collective et, d'autre part, les analyses et courtes citations justifiées par le caractère scientifique ou d'information de l'œuvre dans laquelle elles sont incorporées.

PROJET DE NORME CAMEROUNAISE

ENQUETE PUBLIQUE N° : 18

Durée de l'enquête Du 10/07/2026 Au 08/08 /2026

Edition et diffusion par l'Agence des Normes et de la Qualité

B.P.: 14966 Yaoundé – CAMEROUN – Tél: 6992 257 777 /Fax.: (237) 222 22 64 96

E-mail : enquetepublique@anor.cm – www.anor.cm/enquetes-publiques

ANOR®

DN-R2/PR-01/ENR-02
16/12/2015

INTERNATIONAL
STANDARD

ISO
4214

IDF 254

First edition
2022-12

**Milk and milk products —
Determination of amino acids in infant
and adult/paediatriic nutritional
formulas and dairy products**



Reference numbers
ISO 4214:2022(E)
IDF 254:2022(E)

© ISO and IDF 2022



COPYRIGHT PROTECTED DOCUMENT

© ISO and IDF 2022

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11

Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

International Dairy Federation
Silver Building • Bd Auguste Reyers 70/B
B-1030 Brussels
Phone: +32 2 325 67 40
Fax: +32 2 325 67 41
Email: info@fil-idf.org
Website: www.fil-idf.org

Contents

Page

Forewords	iv
1 Scope	1
2 Normative references	1
3 Terms and definitions	1
4 Principle	1
5 Reagents	2
6 Reagents and standard preparation	3
6.1 General.....	3
6.2 Norvaline (Nva) internal standards.....	4
6.3 Cystine calibration standards.....	4
6.4 Amino acid (AA) calibration standards (with exception of cystine).....	5
6.5 Chromatography solvents (mobile phases).....	5
7 Apparatus	6
8 Sample analysis	6
8.1 Sample preparation.....	6
8.2 Cystine calibration standards preparation.....	7
8.3 Hydrolysis (of samples and cystine standards).....	7
8.4 Neutralization and dilution (of samples and cystine standards).....	7
8.5 Amino acids calibration standards preparation (not needed to be acid hydrolysed).....	8
8.6 Derivatization (of samples, cystine standards and amino acids standards).....	8
8.7 UHPLC separation.....	8
8.8 Peak identification and integration.....	9
9 Calculation and expression of results	10
9.1 Calibration curve.....	10
9.2 Amino acid calculation.....	10
10 Precision	11
10.1 General.....	11
10.2 Repeatability.....	11
10.3 Reproducibility.....	11
11 Test report	15
Annex A (informative) Precision data	16
Annex B (informative) Chromatogram examples	34
Bibliography	35

Forewords

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 34, *Food products*, Subcommittee SC 5, *Milk and milk products*, and the International Dairy Federation (IDF), in collaboration with AOAC INTERNATIONAL and ISO/TC 34, *Food products*, Subcommittee SC 4, *Cereals and pulses*. It is being published jointly by ISO and IDF, and separately by AOAC INTERNATIONAL.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

IDF (the International Dairy Federation) is a non-profit private sector organization representing the interests of various stakeholders in dairying at the global level. IDF members are organized in National Committees, which are national associations composed of representatives of dairy-related national interest groups including dairy farmers, dairy processing industry, dairy suppliers, academics and governments/food control authorities.

ISO and IDF collaborate closely on all matters of standardization relating to methods of analysis and sampling for milk and milk products. Since 2001, ISO and IDF jointly publish their International Standards using the logos and reference numbers of both organizations.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. IDF shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

This document was prepared by the IDF *Standing Committee on Analytical Methods for Composition* and ISO Technical Committee ISO/TC 34, *Food products*, Subcommittee SC 5, *Milk and milk products*, in collaboration with AOAC INTERNATIONAL and ISO/TC 34, *Food products*, Subcommittee SC 4, *Cereals and pulses*. It is being published jointly by ISO and IDF, and separately by AOAC INTERNATIONAL.

The work was carried out by the IDF/ISO Action Team C51 of the *Standing Committee on Analytical Methods for Composition* under the aegis of its project leader Mr C. Fuerer (CH).

Milk and milk products — Determination of amino acids in infant and adult/paediatric nutritional formulas and dairy products

1 Scope

This document specifies a method for the quantitative determination of total amino acids using 6-aminoquinolyl-N-hydroxy-succinimidyl carbamate (ACQ) derivatization followed by ultra-high-performance liquid chromatography (UHPLC) separation and ultraviolet (UV) detection. It specifies a method for the determination, in one single analysis, of the following amino acids: alanine, arginine, aspartic acid (combined with asparagine), cystine (dimer of cysteine, combined with cysteine), glutamic acid (combined with glutamine), glycine, histidine, isoleucine, leucine, lysine, methionine, phenylalanine, proline, serine, threonine, tyrosine and valine.

This method does not apply to the determination of tryptophan.

This method is applicable to infant and adult/paediatric nutritional formulas, dairy products and other matrices such as cereals. It was validated in infant formulas (milk- and soy-based, including partially hydrolysed and elemental products), toddler formula, adult nutritional powder, UHT skimmed milk, whey powder, sodium caseinate, whole milk powder, bran pet food, dry pet food and breakfast cereal (see [Annex A](#) for details).

2 Normative references

There are no normative references in this document.

3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

ISO and IEC maintain terminology databases for use in standardization at the following addresses:

- ISO Online browsing platform: available at <https://www.iso.org/obp>
- IEC Electropedia: available at <https://www.electropedia.org/>

3.1

infant formula

breast-milk substitute specially manufactured to satisfy, by itself, the nutritional requirements of infants during the first months of life up to the introduction of appropriate complementary feeding

[SOURCE: Codex Standard 72-1981]

3.2

adult/paediatric nutritional

nutritionally complete, specially formulated food, consumed in liquid form, which may constitute the sole source of nourishment, made from any combination of milk, soy, rice, whey, hydrolysed protein, starch and amino acids, with and without intact protein

4 Principle

Proteins are hydrolysed in 6 mol/l hydrochloric acid (HCl) for 24 h at 110 °C in the presence of phenol, 3-3'-dithiodipropionic acid (DDP) and norvaline. Phenol is added to prevent halogenation of tyrosine.

Norvaline is added as an internal standard. DDP is added to convert cystine and cysteine to S-2-carboxyethylthiocysteine (XCys), as described in Reference [1], and the resulting derivative can be separated from other amino acids for quantification.

After neutralization, amino acids and converted XCys are derivatized with AQC. Derivatized amino acids are separated using reversed phase UHPLC with UV detection at 260 nm.

NOTE Fluorescence detection can be used provided equivalence has been demonstrated.

During acid hydrolysis, glutamine (Gln) and asparagine (Asn) are converted to glutamic acid (Glu) and aspartic acid (Asp), respectively. Thus, Glu values represent the combined values of Glu and Gln, and Asp values represent the combined values of Asp and Asn. Cys2 values represent the combined values of cysteine and cystine since both are converted to XCys by DDP.

5 Reagents

Use only reagents of recognized analytical grade.

Commercial references are only a guideline. Equivalent chemicals or materials can be used provided their equivalence has been demonstrated. Before using chemicals, refer to the safety data sheets and ensure that the safety precautions are applied.

5.1 AccQ·Tag™ Ultra Derivatization Kit (Waters 186003836¹⁾).

As an alternative derivatizing buffer, di-sodium tetraborate decahydrate (CAS Registry Number®²⁾ 1303-96-4) can be used.

As an alternative tagging reagent, 6-aminoquinolyl-N-hydroxysuccinimidyl carbamate (CAS 148757-94-2) can be used.

5.2 AccQ·Tag™ Ultra Eluent A concentrate (Waters 186003838¹⁾).

As alternative reagents, acetonitrile, gradient grade for liquid chromatography (LC) (CAS 75-05-8), and formic acid (CAS 64-18-6) can be used.

5.3 AccQ·Tag™ Ultra Eluent B (Waters 186003839¹⁾).

As alternative reagents, acetonitrile, gradient grade for LC (CAS 75-05-8), formic acid (CAS 64-18-6) and ammonium formate (CAS 540-69-2) can be used.

5.4 Phenol (CAS 108-95-2).

5.5 3,3'-Dithiodipropionic acid (CAS 1119-62-6).

5.6 Amino acid standard solution, containing the following 17 amino acids at 2,5 µmol/ml each (except L-cystine at 1,25 µmol/ml) in 0,1 mol/l HCl: L-alanine, L-arginine, L-aspartic acid, L-cystine, L-glutamic acid, L-glycine, L-histidine, L-isoleucine, L-leucine, L-lysine, L-methionine, L-phenylalanine, L-proline, L-serine, L-threonine, L-tyrosine and L-valine.

5.7 L-cystine (CAS 56-89-3).

1) This is an example of a suitable product available commercially. This information is given for the convenience of users of this document and does not constitute an endorsement by ISO or IDF of the product named. The alternative reagents listed in this document have been shown to lead to the same results.

2) CAS Registry Number® is a trademark of CAS corporation. This information is given for the convenience of users of this document and does not constitute an endorsement by ISO of the product named. Equivalent products may be used if they can be shown to lead to the same results.

5.8 Norvaline (CAS 6600-40-4).

5.9 Sodium hydroxide pellets, reagent grade (CAS 1310-73-2).

5.10 Sodium hydroxide solution (CAS 1310-73-2), substance concentration $c = 1$ mol/l.

5.11 Sodium hydroxide solution (optional) (CAS 1310-73-2), $c = 6$ mol/l.

5.12 Hydrochloric acid fuming 37 % (CAS 7647-01-0), $c = 12$ mol/l, GR grade for analysis.

5.13 Hydrochloric acid solution (CAS 7647-01-0), $c = 1$ mol/l.

5.14 Hydrochloric acid solution (CAS 7647-01-0), $c = 0,1$ mol/l.

5.15 Laboratory water, with a resistivity of 18,2 M Ω -cm (ultra-pure water).

6 Reagents and standard preparation

6.1 General

6.1.1 Sodium hydroxide (NaOH) solutions, $c = 6$ mol/l, $c = 0,2$ mol/l and $c = 0,05$ mol/l.

For the $c = 6$ mol/l solution, weigh out 24 g of sodium hydroxide (5.9) into a 100 ml volumetric flask. Dissolve in about 80 ml of water. Allow to cool down and dilute to the mark with water. Optional: use a commercially available equivalent (5.11).

For the $c = 0,2$ mol/l solution, pipet 20 ml of 1 mol/l NaOH (5.10) into a 100 ml volumetric flask and make up to the mark with water.

For the $c = 0,05$ mol/l solution, pipet 5 ml of 1 mol/l NaOH (5.10) into a 100 ml volumetric flask and make up to the mark with water.

6.1.2 Hydrochloric acid (HCl) solution, $c = 0,2$ mol/l.

Pipet 20 ml of 1 mol/l HCl (5.13) into a 100 ml volumetric flask and make up to the mark with water.

6.1.3 DDP solution, mass concentration $\rho = 10$ g/l (in NaOH, $c = 0,2$ mol/l).

Into a 50 ml volumetric flask, weigh out 500 mg of DDP and make up to the mark with 0,2 mol/l NaOH (6.1.1).

6.1.4 Phenol/HCl solution, $\rho = 1$ g/l (in HCl, $c = 12$ mol/l).

Into a 100 ml volumetric flask, weigh out 100 mg of phenol and make up to the mark with 12 mol/l HCl (5.12).

6.1.5 AccQ·Tag™ Ultra Derivatization kit.¹⁾

Prepare the reagents included in the kit following the manufacturer's instructions.

6.1.6 AccQ·Tag™ Ultra Borate buffer (reagent 1).¹⁾

Ready-to-use solution. An alternative reagent is sodium tetraborate in water solution ($\rho = 50$ g/l). Into a 100 ml volumetric flask, weigh 5 g of sodium tetraborate decahydrate, dissolve and make up to the mark with water.

6.1.7 AccQ·Tag™ Ultra reagent (vial 2A and 2B).¹⁾

Reconstitute AccQ·Tag™ Ultra reagent (vial 2A) according to the manufacturer's instructions as follows:

- a) Preheat a heating block to 55 °C.
- b) Tap vial 2A lightly before opening to ensure all AccQ·Tag™ Ultra reagent powder is at the bottom of the vial.
- c) Rinse a clean micropipette by drawing and discarding 1 ml of AccQ·Tag™ Ultra reagent diluent from vial 2B (ready-to-use solution). Repeat twice.
- d) Draw 1,0 ml from vial 2B and transfer it to the AccQ·Tag™ Ultra reagent powder in vial 2A. Cap the vial tightly.
- e) Vortex mix for approximately 10 s.
- f) Heat vial 2A on top of the preheated heating block until the AccQ·Tag™ Ultra reagent powder is dissolved. Do not heat the reagent for longer than 10 min.

Once reconstituted, the AccQ·Tag™ Ultra reagent concentration is approximately 10 mmol/l. Store the reconstituted AccQ·Tag™ Ultra reagent in a desiccator at room temperature for up to one week.

CAUTION — AccQ·Tag™ Ultra reagent reacts with atmospheric moisture. Seal the container tightly when not in use. Do not refrigerate. Do not use discoloured reagent, especially if it is yellow or green.

The following alternative reagent can be used. Into a 4 ml vial, weigh out approximately 3,0 mg to 4,0 mg of 6-aminoquinolyl-N-hydroxysuccinimidyl carbamate. Continue with step c) above using LC-grade acetonitrile instead of the AccQ·Tag™ Ultra reagent diluent.

6.2 Norvaline (Nva) internal standards

6.2.1 Nva stock solution, $c = 10$ mmol/l.

Weigh 117,16 mg Nva into a 100 ml volumetric flask and make up to the mark with 0,1 mol/l HCl ([5.14](#)).

6.2.2 Nva solution, $c = 2,5$ mmol/l.

Pipet 2,5 ml of Nva stock solution ([6.2.1](#)) into a 10 ml volumetric flask and make up to the mark with 0,1 mol/l HCl ([5.14](#)).

Store both Nva solutions at -20 °C for up to six months as 2 ml aliquots.

6.3 Cystine calibration standards

6.3.1 Cystine stock solution, $c = 10$ mmol/l.

Weigh 240 mg cystine into a 100 ml volumetric flask and make up to the mark with 0,05 mol/l NaOH ([6.1.1](#)). Store this solution at -20 °C for up to three months as 1 ml aliquots.

6.3.2 Cystine solution, $c = 1$ mmol/l.

Add 900 µl of 0,05 mol/l NaOH ([6.1.1](#)) to 100 µl of cystine stock solution ([6.3.1](#)). Prepare this solution freshly for each analysis.

6.4 Amino acid (AA) calibration standards (with exception of cystine)

6.4.1 AA stock solution, $c = 2,5$ mmol/l.

Amino acid standard solution is ready-to-use and contains 2,5 mmol/l of each amino acid (although present in this solution, cystine is not used for quantification and is prepared separately, see [6.3](#)).

Store this calibration standard stock solution at -20 °C for up to six months as 250 µl aliquots.

6.4.2 AA solution 1, $c = 0,5$ mmol/l.

Add 600 µl of 0,1 mol/l HCl ([5.14](#)) to 150 µl of AA stock solution ([6.4.1](#)). Prepare this solution freshly for each analysis.

6.4.3 AA solution 2, $c = 0,05$ mmol/l.

Add 900 µl of 0,1 mol/l HCl ([5.14](#)) to 100 µl of AA solution 1 ([6.4.2](#)). Prepare this solution freshly for each analysis.

6.5 Chromatography solvents (mobile phases)

6.5.1 Eluent A (Solvent A).

Prepare Eluent A from AccQ·Tag™ Ultra Eluent A concentrate as follows:

- Measure 850 ml of water into a 1 l graduated cylinder.
- In a separate graduated cylinder, measure 150 ml of AccQ·Tag™ Ultra Eluent A concentrate.
- Add the concentrate to the water and mix thoroughly.

Eluent A concentrate, once opened, shall be stored tightly capped at around 4 °C. Dilute Eluent A is stable for one week at room temperature.

As an alternative for Eluent A concentrate, the following solution can be used. Mix 840 ml of 200 mmol/l ammonium formate solution (12,61 g ammonium formate in 1 l of water), 110 ml of acetonitrile and 50 ml of formic acid. Prepare Eluent A from the concentrate as described above.

6.5.2 Eluent B (Solvent B).

AccQ·Tag™ Eluent B is supplied as a working solution. No additional preparation is required. Eluent B, once opened, shall be stored tightly capped at around 4 °C for no longer than one month.

As an alternative for Eluent B, the following solution can be used. Add 13,2 ml formic acid to 1 l acetonitrile.

6.5.3 Wash solvents.

- The weak needle wash solvent is 50 ml/l acetonitrile in water.
- The strong needle wash solvent is 950 ml/l acetonitrile in water.
- The seal wash solvent is 500 ml/l acetonitrile in water.

7 Apparatus

7.1 UHPLC system that sustains a pressure of approximately 62 MPa and can achieve baseline separation of the amino acids³⁾.

7.2 Chromatography column, ACQUITY UPLC™ BEH C18 Column, 130 Å, 1,7 µm, 2,1 mm × 150 mm (Waters 186002353¹⁾) or equivalent, provided baseline separation of the amino acids is achieved.

7.3 Adjustable micropipettes, of volume 10 µl, 20 µl, 200 µl and 1 000 µl and tips.

7.4 Vortex mixer.

7.5 Analytical balance, with a precision of 0,1 mg.

7.6 Heating block, able to maintain a temperature of 55 °C ± 2 °C.

7.7 Laboratory oven, able to maintain a temperature of 110 °C ± 2 °C.

7.8 Syringe filter, 0,45 µm Polyvinylidene fluoride (PVDF) Millex®-HV (e.g. Millipore SLHV013NL¹⁾ or equivalent).

7.9 Syringes, of 2 ml volume.

7.10 Borosilicate glass tubes (e.g. Pyrex), of 10 ml volume with screw cap.

7.11 Microtubes, of volume 1,5 ml and 2 ml.

7.12 Vial with screw cap, of volume 4 ml.

7.12.1 Glass screw neck total recovery vial, 12 mm × 32 mm (Waters 186000384C¹⁾ or equivalent).

8 Sample analysis

8.1 Sample preparation

Prepare different sample types differently. Powder infant and adult/paediatric nutritional formulas shall be reconstituted in water first. Other samples are used without reconstitution.

Reconstitute powder infant and adult/paediatric nutritional formula samples by adding 25 g powder to 200 g water and mix thoroughly. Weigh 220 mg ± 20 mg reconstituted powders into a 10 ml glass tube with screw cap. Report the sample mass to 0,1 mg.

For ready-to-feed infant and adult/paediatric nutritional formula and liquid dairy samples, weigh 220 mg ± 20 mg liquid into a 10 ml glass tube with screw cap. Report the sample mass to 0,1 mg.

For dairy powder and cereals samples, weigh 50 mg ± 5 mg dairy powder samples or 100 mg ± 10 mg cereals samples into a 10 ml glass tube with screw cap. Report the sample mass to 0,1 mg.

To each tube, add water, DDP, HCl, Nva and phenol/HCl according to [Table 1](#).

3) Agilent 1260 Infinity II¹⁾, Waters Acquity¹⁾, Waters Acquity-I¹⁾, Waters Acquity-H¹⁾ and Thermo Fisher Scientific 3000 RS¹⁾ have been successfully used during the multi-laboratory study.

Add the phenol/HCl solution under the hood. Sparge the tube a minimum of approximately 5 s with a stream of nitrogen to displace oxygen. Close tubes with screw caps and vortex. Make sure the caps are perfectly clean (i.e. devoid of any particle) to ensure tightness and avoid evaporation during hydrolysis.

Table 1 — Preparation of the sample tubes

Solution	Reconstituted infant formulas and adult nutritionals, ready-to-feed formulas and liquid dairy samples	Dairy powders	Cereals
Sample, mg	220	50	100
Water, µl	880	750	1 000
DDP solution (6.1.3), µl	600	600	600
0,2 mol/l HCl (6.1.2), µl	600	600	600
Nva stock solution (6.2.1), µl	200	500	200
Phenol/HCl solution (6.1.4), µl	2 500	2 500	2 500

8.2 Cystine calibration standards preparation

Table 2 describes how to prepare calibration standards for converted cystine at 0 pmol/µl to 10 pmol/µl with Nva at 10 pmol/µl (all are final concentrations after derivatization).

Add the phenol/HCl solution under the hood. Sparge the tube approximately 5 s with a stream of nitrogen to displace oxygen. Close tubes with screw caps and vortex. Make sure the caps are perfectly clean (i.e. devoid of any particle) to ensure tightness and avoid evaporation during hydrolysis.

Table 2 — Final cystine concentration after derivatization

Solution	10 pmol/µl	5 pmol/µl	2,5 pmol/µl	1 pmol/µl	0,5 pmol/µl	0 pmol/µl
Cystine solution, µl	200 ^a	100 ^a	50 ^a	200 ^b	100 ^b	0
Water, µl	900	1 000	1 050	900	1 000	1 100
DDP solution (6.1.3), µl	600	600	600	600	600	600
0,2 mol/l HCl (6.1.2), µl	600	600	600	600	600	600
Nva stock solution (6.2.1), µl	200	200	200	200	200	200
Phenol/HCl solution (6.1.4), µl	2 500	2 500	2 500	2 500	2 500	2 500
^a 10 mmol/l cystine stock solution (6.3.1).						
^b 1 mmol/l cystine solution (6.3.2).						

8.3 Hydrolysis (of samples and cystine standards)

Place tubes in an oven at 110 °C ± 2 °C for 24 h ± 0,5 h.

8.4 Neutralization and dilution (of samples and cystine standards)

Take the tubes out of the oven. Allow hydrolysates to cool down and particles to settle down prior to taking an aliquot. When transferring aliquots, pipet about 1 cm below the top of the liquid. Perform neutralization under the hood.

For infant formula, liquid dairy and cereal samples as well as converted cystine standards, transfer 0,2 ml of each hydrolysate (samples and converted cystine standards) into a 1,5 ml microtube, add 0,2 ml of 6 mol/l NaOH (6.1.1) and then 0,4 ml of 0,1 mol/l HCl (5.14). Mix well and filter through a 0,45 µm membrane filter into another 1,5 ml microtube.

For dairy powder samples, transfer 0,2 ml of the hydrolysate sample solution into a 2 ml microtube, add 0,2 ml of 6 mol/l NaOH (6.1.1) and then 1,6 ml of 0,1 mol/l HCl (5.14). Mix well and filter through 0,45 µm membrane filter into another 2 ml microtube.

8.5 Amino acids calibration standards preparation (not needed to be acid hydrolysed)

Table 3 shows how to prepare 0,5 ml calibration standards at 0 pmol/µl to 25 pmol/µl and Nva at 10 pmol/µl (all are final concentration after derivatization).

The amino acid solutions are stable for one week when stored at 4 °C ± 2 °C.

Table 3 — Amino acid concentrations (each, final, after derivatization)

Solution	25 pmol/µl	10 pmol/µl	5 pmol/µl	1 pmol/µl	0,5 pmol/µl	0 pmol/µl
Amino acid solution, µl	50 ^a	100 ^b	50 ^b	100 ^c	50 ^c	0
Water, µl	50	100	50	100	50	0
Nva solution (6.2.2), µl	20	20	20	20	20	20
0,1 mol/l HCl (5.14), µl	380	280	380	280	380	480
^a 2,5 mmol/l AA stock solution (6.4.1).						
^b 0,5 mmol/l AA solution 1 (6.4.2).						
^c 0,05 mmol/l AA solution 2 (6.4.3).						

8.6 Derivatization (of samples, cystine standards and amino acids standards)

Derivatization converts free amino acids into highly stable derivatives. Standards and samples are derivatized following the manufacturer's instructions, as follows:

- Preheat a heating block to 55 °C.
- With a micropipette, add 70 µl of AccQ·Tag™ Ultra Borate buffer (reagent 1, see 6.1.6) to a clean 12 mm × 32 mm glass screw neck total recovery vial.
- Add 10 µl of calibration standard (8.5), neutralized sample solution (8.4), or neutralized converted cystine standard (8.4) to the vial.
- Vortex mix briefly.
- Add 20 µl of reconstituted AccQ·Tag™ Ultra reagent (6.1.7) to the sample vial.
- Mix the solution immediately by pipetting up and down several times. Cap and mix by vortex immediately for several seconds and tap the vial to ensure no bubble is trapped.
- Let stand for 1 min at room temperature.
- Heat the vial in a heating block for 10 min at 55 °C ± 1 °C.

8.7 UHPLC separation

Perform UHPLC separation as follows:

- Prime solvent lines for 5 min.
- Prime wash/sample syringes for four cycles.
- Allow the chromatographic system to stabilize before injecting standards and samples. Make sure the system pressure and initial conditions are stable before performing injections (around 62 MPa).
- Before starting a series of analyses, inject two blanks (water) to condition the column.

- Inject 1 µl of each derivatized calibration standards, and then inject 1 µl of derivatized sample solutions. Perform single injections. Add a blank injection (water) at the end of each calibration series.
- Perform UHPLC under the following conditions and those given in [Table 4](#):
 - column temperature: 50 °C;
 - UV detector: 260 nm;
 - injection volume: 1 µl;
 - flow rate: 0,4 ml/min;
 - mobile phase A: Eluent A ([6.5.1](#));
 - mobile phase B: Eluent B ([6.5.2](#)).

NOTE Fluorescence detection can be used provided equivalence has been demonstrated.

Operating conditions can vary depending on the apparatus. Follow the supplier's instructions. Examples of wash solvents used with Waters UPLC systems¹⁾ are given in [6.5.3](#).

Table 4 — Eluent gradient

Time	% A	% B	Curve
0,00	99,9	0,1	
5,50	99,9	0,1	2
15,22	90,9	9,1	7
20,47	78,8	21,2	6
21,26	40,4	59,6	6
21,29	10	90	6
22,84	10	90	6
26,00	99,9	0,1	6
32,00	99,9	0,1	6

8.8 Peak identification and integration

Identify the amino acids peaks in the sample solution by comparison with the retention times of the corresponding peaks obtained in the calibration standards (see [Annex B](#) for examples). If a peak has not been integrated correctly, call the recorded data and reintegrate.

To verify system stability, inject a mid-level standard a minimum of three times (5 × for United States Pharmacopoeia (USP) requirements) and ensure response and retention times have a coefficient of variation, $C_v < 2$.

Check that peaks are separated with a good resolution (baseline separation). If this is not the case, adapt the chromatographic conditions (gradient, temperature, tubing length, etc.) accordingly.

To check that the derivatization reagent was present in sufficient amount (excess), verify that the derivatization peak is present in the chromatogram. The response of the excess reagent is present in the chromatography as the large 6-aminoquinoline (AMQ) peak first to elute. The response should be equal to that of the 25 pmol/µl standard or the reaction and sample should be flagged and discarded. The derivatization peak at approximately 17 min prior to lysine can be ignored.

9 Calculation and expression of results

9.1 Calibration curve

Establish the calibration curve from the six different calibration standards for each amino acid and converted cystine at the beginning of each series of analyses by plotting the response (peak area ratio of analyte versus internal standard, multiplied by the concentration of the internal standard, see [Formula \(1\)](#) against analyte concentration:

$$R = \frac{A_s}{A_{is}} \times c_{is} \quad (1)$$

where

R is the response;

A_s is the peak area of individual amino acids in the calibration standard solution;

A_{is} is the peak area of internal standard in the calibration standard solution;

c_{is} is the concentration of internal standard in the calibration standard solution, in pmol/μl.

Force the linear regression through zero. Check the linearity of the calibration (the correlation coefficient R^2 shall be above 0,99).

9.2 Amino acid calculation

Calculate the amount of individual amino acids present in the sample in pmol/μl from the calibration curve using [Formula \(2\)](#):

$$c_s = \frac{A_s \times c_{is}}{A_{is} \times S} \quad (2)$$

where

c_s is the concentration of individual amino acid in the test sample solution in pmol/μl;

A_s is the peak area of individual amino acid in the test sample solution;

c_{is} is the concentration of internal standard in the test sample solution, in pmol/μl;

A_{is} is the peak area of internal standard in the test sample solution;

S is the slope of the calibration curve (all curves are forced through zero, equation: $y = ax$).

Calculate the mass fraction, w , of each amino acid, in milligram per 100 g of product, using [Formula \(3\)](#):

$$w = \frac{c_s \times M_A \times V_s \times d_1 \times d_2}{m_s \times 10} \quad (3)$$

where

c_s is the concentration of individual amino acid in the test sample solution in pmol/μl;

M_A is the molar mass of individual amino acids in g/mol (see [Table 5](#));

V_s is the volume of hydrolysis solution in ml (typically $V_s = 5$ ml);

d_1 is the dilution factor in the neutralization step (4 or 10, depending on the sample);

- d_2 is the dilution factor in the derivatization step (10);
- m_s is the mass of the test portion in mg;
- 10 is the combined factor to convert pg to mg (10^{-9}), ml to μl (10^3) and mg to 100 g ($1/10^{-5}$).

Table 5 — Molar mass (M_A) of amino acids

Amino acid	M_A g/mol	Amino acid	M_A g/mol
Alanine	89,10	Lysine	146,19
Arginine	174,20	Methionine	149,21
Aspartic acid	133,11	Phenylalanine	165,19
Cystine	240,30	Proline	115,13
Glutamic acid	147,13	Serine	105,09
Glycine	75,07	Threonine	119,12
Histidine	155,16	Tyrosine	181,19
Isoleucine	131,18	Valine	117,15
Leucine	131,18		

10 Precision

10.1 General

Details of the interlaboratory test of the precision of the method are summarized in [Annex A](#). The values derived from the interlaboratory test may not be applicable to analyte concentration ranges and/or matrices other than those given in [Annex A](#).

10.2 Repeatability

The absolute difference between two single test results found on identical test material by one operator using the same apparatus with the shortest feasible time interval will exceed the repeatability limit r in not more than 5 % of the cases. The values of r are given in [Table 6](#).

10.3 Reproducibility

The absolute difference between two single test results found on identical test material reported by two laboratories will exceed the reproducibility limit R in not more than 5 % of the cases. The values of R are given in [Table 6](#).

Table 6 — Precision data

Alanine																
Sample	S1 ^a	S2 ^b	S3 ^c	S4 ^d	S5 ^e	S6 ^f	S7 ^g	S8 ^h	S9 ⁱ	D1 ^j	D2 ^k	D3 ^l	D4 ^m	C1 ⁿ	C2 ^o	C3 ^p
$\bar{x}$, mg/100 g	60,0	66,5	59,9	60,9	59,7	49,4	60,9	59,7	86,7	107,8	516,1	2 756,4	816,7	421,5	756,4	271,5
r , mg/100 g	3,94	2,21	2,91	2,41	2,70	1,46	2,57	3,20	3,59	7,52	38,03	108,16	46,37	13,37	53,87	15,64
R , mg/100 g	14,93	8,47	7,60	9,31	6,96	5,79	8,48	7,47	13,56	16,23	77,36	546,48	123,27	52,01	113,78	30,31
Arginine																
Sample	S1 ^a	S2 ^b	S3 ^c	S4 ^d	S5 ^e	S6 ^f	S7 ^g	S8 ^h	S9 ⁱ	D1 ^j	D2 ^k	D3 ^l	D4 ^m	C1 ⁿ	C2 ^o	C3 ^p
$\bar{x}$, mg/100 g	62,8	32,7	102,3	106,4	32,9	36,5	103,5	50,6	89,1	115,0	256,4	3 448,9	850,7	463,3	839,1	254,2
r , mg/100 g	6,44	2,42	5,26	5,99	4,12	3,14	6,12	2,48	6,76	12,12	52,87	171,83	51,03	27,21	104,91	19,54
R , mg/100 g	14,45	4,32	13,80	11,39	5,16	5,82	13,28	5,89	10,06	27,48	52,87	565,90	91,17	91,90	135,54	58,31
Asparagine and aspartic acid																
Sample	S1 ^a	S2 ^b	S3 ^c	S4 ^d	S5 ^e	S6 ^f	S7 ^g	S8 ^h	S9 ⁱ	D1 ^j	D2 ^k	D3 ^l	D4 ^m	C1 ⁿ	C2 ^o	C3 ^p
$\bar{x}$, mg/100 g	143,1	150,7	163,8	182,0	136,9	112,8	166,8	139,0	137,2	259,3	1 148,0	6 380,1	1 950,0	619,6	1 791,7	374,4
r , mg/100 g	3,47	11,87	11,46	25,22	8,70	3,98	12,80	7,34	6,86	19,66	67,31	324,55	173,10	25,38	152,94	28,05
R , mg/100 g	19,46	17,61	18,03	25,86	26,58	13,14	21,86	14,80	21,25	64,90	155,39	1 540,26	253,03	86,73	370,90	60,48
Cysteine and cystine																
Sample	S1 ^a	S2 ^b	S3 ^c	S4 ^d	S5 ^e	S6 ^f	S7 ^g	S8 ^h	S9 ⁱ	D1 ^j	D2 ^k	D3 ^l	D4 ^m	C1 ⁿ	C2 ^o	C3 ^p
$\bar{x}$, mg/100 g	18,1	35,1	18,0	24,4	20,3	19,6	18,0	20,9	16,7	25,9	258,9	389,1	199,8	177,2	187,8	125,6
r , mg/100 g	2,73	1,67	1,37	3,95	2,71	1,19	2,31	0,74	0,84	2,06	12,40	27,93	19,63	15,70	20,60	8,10
R , mg/100 g	6,95	11,20	5,57	8,17	5,70	6,46	6,24	3,80	7,24	5,13	47,63	118,68	21,50	48,67	30,66	30,33
Glutamine and glutamic acid																
Sample	S1 ^a	S2 ^b	S3 ^c	S4 ^d	S5 ^e	S6 ^f	S7 ^g	S8 ^h	S9 ⁱ	D1 ^j	D2 ^k	D3 ^l	D4 ^m	C1 ⁿ	C2 ^o	C3 ^p
$\bar{x}$, mg/100 g	336,7	242,4	276,8	234,7	279,3	248,4	282,9	334,3	242,7	720,0	1 949,6	20 287,3	5 389,8	2 663,7	2 313,9	2 010,7
r , mg/100 g	7,05	9,68	15,03	6,27	11,88	6,18	9,80	17,82	10,84	48,54	186,89	930,50	311,87	93,90	241,76	121,51
R , mg/100 g	33,74	24,42	34,03	23,87	27,81	18,87	23,30	30,25	26,81	136,04	225,89	5 174,93	684,89	370,85	376,24	290,40
Glycine																
Sample	S1 ^a	S2 ^b	S3 ^c	S4 ^d	S5 ^e	S6 ^f	S7 ^g	S8 ^h	S9 ⁱ	D1 ^j	D2 ^k	D3 ^l	D4 ^m	C1 ⁿ	C2 ^o	C3 ^p
$\bar{x}$, mg/100 g	36,2	24,7	57,9	46,8	27,4	22,6	58,3	30,6	150,7	62,2	228,0	1 745,0	471,3	495,7	789,3	313,0
r , mg/100 g	1,61	0,88	3,43	4,27	1,81	0,97	3,76	1,91	6,28	7,50	36,65	122,58	40,57	21,36	22,27	22,24
R , mg/100 g	8,30	4,15	7,19	7,51	3,14	3,46	9,04	4,79	25,29	9,16	48,52	327,30	59,32	84,88	66,42	48,49

Table 6 (continued)

Histidine																
Sample	S1 ^a	S2 ^b	S3 ^c	S4 ^d	S5 ^e	S6 ^f	S7 ^g	S8 ^h	S9 ⁱ	D1 ^j	D2 ^k	D3 ^l	D4 ^m	C1 ⁿ	C2 ^o	C3 ^p
$\bar{x}$, mg/100 g	40,3	24,0	35,1	44,0	24,8	27,2	34,8	40,2	43,9	88,5	187,2	2 706,5	668,5	241,8	365,3	147,8
<i>r</i> , mg/100 g	1,93	0,47	5,93	2,45	1,36	1,29	3,01	0,95	1,62	3,30	17,08	230,24	37,49	16,95	22,98	14,35
<i>R</i> , mg/100 g	10,38	4,46	8,32	5,99	2,76	6,28	5,31	7,72	5,26	9,42	32,98	600,16	74,10	46,97	40,51	29,75
Isoleucine																
Sample	S1 ^a	S2 ^b	S3 ^c	S4 ^d	S5 ^e	S6 ^f	S7 ^g	S8 ^h	S9 ⁱ	D1 ^j	D2 ^k	D3 ^l	D4 ^m	C1 ⁿ	C2 ^o	C3 ^p
$\bar{x}$, mg/100 g	86,0	84,0	66,8	120,9	83,3	69,6	68,1	88,5	87,1	175,9	670,6	4 854,9	1 296,0	352,4	657,6	254,1
<i>r</i> , mg/100 g	2,47	4,04	4,70	4,15	3,27	2,83	5,36	4,61	1,69	10,78	52,94	169,94	52,39	20,28	57,87	21,80
<i>R</i> , mg/100 g	9,45	9,83	8,75	11,59	12,64	7,65	7,00	8,96	7,01	36,10	52,94	772,90	199,66	63,44	96,38	55,89
Leucine																
Sample	S1 ^a	S2 ^b	S3 ^c	S4 ^d	S5 ^e	S6 ^f	S7 ^g	S8 ^h	S9 ⁱ	D1 ^j	D2 ^k	D3 ^l	D4 ^m	C1 ⁿ	C2 ^o	C3 ^p
$\bar{x}$, mg/100 g	153,0	141,3	110,9	197,0	147,4	124,7	113,4	161,6	138,0	331,3	1 104,1	8 660,8	2 424,4	672,9	1 151,8	484,8
<i>r</i> , mg/100 g	5,09	3,74	6,26	6,71	2,18	4,58	9,50	7,66	3,61	7,87	85,20	286,78	103,00	13,98	84,69	26,57
<i>R</i> , mg/100 g	14,57	12,14	8,21	18,16	16,86	10,96	10,58	12,12	9,29	31,93	100,36	1 016,61	239,93	72,48	102,61	60,96
Lysine																
Sample	S1 ^a	S2 ^b	S3 ^c	S4 ^d	S5 ^e	S6 ^f	S7 ^g	S8 ^h	S9 ⁱ	D1 ^j	D2 ^k	D3 ^l	D4 ^m	C1 ⁿ	C2 ^o	C3 ^p
$\bar{x}$, mg/100 g	131,4	122,9	86,1	110,2	97,4	99,5	84,2	130,7	91,1	260,1	959,1	7 210,1	1 953,0	190,9	813,2	99,5
<i>r</i> , mg/100 g	7,56	8,51	5,93	13,54	4,58	3,89	6,49	11,42	5,27	33,99	111,63	527,39	152,28	16,90	83,12	15,21
<i>R</i> , mg/100 g	30,40	12,61	10,70	14,48	25,69	11,16	10,23	15,71	13,50	64,10	119,22	1 580,88	292,99	39,40	98,12	26,74
Methionine																
Sample	S1 ^a	S2 ^b	S3 ^c	S4 ^d	S5 ^e	S6 ^f	S7 ^g	S8 ^h	S9 ⁱ	D1 ^j	D2 ^k	D3 ^l	D4 ^m	C1 ⁿ	C2 ^o	C3 ^p
$\bar{x}$, mg/100 g	51,6	27,9	36,6	43,7	30,2	29,0	42,1	39,9	37,3	87,5	214,9	2 655,2	643,3	136,1	138,3	101,9
<i>r</i> , mg/100 g	1,65	3,17	2,46	2,87	2,78	1,67	1,17	1,96	1,11	11,35	29,05	161,65	51,50	8,05	22,97	9,30
<i>R</i> , mg/100 g	8,46	5,67	4,71	5,44	4,97	3,42	5,08	5,48	4,63	21,04	69,45	440,70	115,51	31,98	74,14	20,12
Phenylalanine																
Sample	S1 ^a	S2 ^b	S3 ^c	S4 ^d	S5 ^e	S6 ^f	S7 ^g	S8 ^h	S9 ⁱ	D1 ^j	D2 ^k	D3 ^l	D4 ^m	C1 ⁿ	C2 ^o	C3 ^p
$\bar{x}$, mg/100 g	73,5	42,9	72,9	100,4	57,4	50,5	73,6	71,4	76,5	164,4	341,4	4 856,9	1 212,2	457,3	735,9	323,0
<i>r</i> , mg/100 g	2,58	4,48	2,93	6,94	5,19	4,21	7,21	4,48	6,27	5,74	13,47	538,26	79,83	41,50	54,43	24,08
<i>R</i> , mg/100 g	5,84	6,88	8,62	8,54	9,12	6,02	7,21	9,05	9,20	15,19	53,37	1 085,91	146,31	74,15	89,31	49,79

Table 6 (continued)

Proline																
Sample	S1 ^a	S2 ^b	S3 ^c	S4 ^d	S5 ^e	S6 ^f	S7 ^g	S8 ^h	S9 ⁱ	D1 ^j	D2 ^k	D3 ^l	D4 ^m	C1 ⁿ	C2 ^o	C3 ^p
$\bar{x}$, mg/100 g	136,2	79,1	71,8	60,4	113,8	100,4	73,3	140,7	128,5	334,0	635,4	9 892,8	2 404,1	859,0	902,3	671,9
<i>r</i> , mg/100 g	4,27	2,71	3,23	3,87	5,57	2,52	4,74	3,98	6,56	7,22	30,99	315,21	105,69	29,14	78,04	55,70
<i>R</i> , mg/100 g	12,47	7,98	4,19	6,36	14,63	8,73	6,69	11,58	10,99	35,67	61,81	1 307,39	245,37	78,45	83,65	99,82
Serine																
Sample	S1 ^a	S2 ^b	S3 ^c	S4 ^d	S5 ^e	S6 ^f	S7 ^g	S8 ^h	S9 ⁱ	D1 ^j	D2 ^k	D3 ^l	D4 ^m	C1 ⁿ	C2 ^o	C3 ^p
$\bar{x}$, mg/100 g	89,6	72,1	74,8	56,0	78,9	67,9	75,8	89,3	70,4	190,4	554,8	5 378,2	1 400,8	471,4	780,5	335,9
<i>r</i> , mg/100 g	4,91	3,05	5,49	3,12	3,44	2,44	5,16	4,46	2,37	7,03	36,60	218,35	82,09	24,46	43,45	19,46
<i>R</i> , mg/100 g	16,09	11,49	13,61	9,94	11,56	10,75	12,98	13,57	10,52	35,98	142,32	1 242,03	245,78	106,04	75,43	52,40
Threonine																
Sample	S1 ^a	S2 ^b	S3 ^c	S4 ^d	S5 ^e	S6 ^f	S7 ^g	S8 ^h	S9 ⁱ	D1 ^j	D2 ^k	D3 ^l	D4 ^m	C1 ⁿ	C2 ^o	C3 ^p
$\bar{x}$, mg/100 g	76,9	95,1	53,7	77,7	86,1	67,5	55,1	81,8	68,9	150,9	719,3	4 048,8	1 099,2	319,9	611,6	217,2
<i>r</i> , mg/100 g	1,52	2,29	2,22	5,59	4,15	1,73	2,03	3,94	1,95	4,32	34,59	154,91	59,51	13,79	32,58	14,18
<i>R</i> , mg/100 g	11,28	10,41	6,80	10,21	11,17	8,10	6,50	8,91	7,77	17,06	87,67	686,35	127,47	33,67	50,66	26,08
Tyrosine																
Sample	S1 ^a	S2 ^b	S3 ^c	S4 ^d	S5 ^e	S6 ^f	S7 ^g	S8 ^h	S9 ⁱ	D1 ^j	D2 ^k	D3 ^l	D4 ^m	C1 ⁿ	C2 ^o	C3 ^p
$\bar{x}$, mg/100 g	72,8	38,0	51,7	92,3	55,4	50,1	53,0	73,1	66,9	172,2	289,4	5 481,0	1 229,6	272,8	478,8	192,2
<i>r</i> , mg/100 g	1,76	1,19	5,18	8,22	5,88	2,47	6,53	4,36	5,32	4,00	61,07	546,42	92,01	32,69	29,42	14,54
<i>R</i> , mg/100 g	10,37	5,91	6,87	12,38	8,02	8,63	8,97	8,74	8,01	16,98	85,64	1 077,74	198,74	96,37	65,82	20,51
Valine																
Sample	S1 ^a	S2 ^b	S3 ^c	S4 ^d	S5 ^e	S6 ^f	S7 ^g	S8 ^h	S9 ⁱ	D1 ^j	D2 ^k	D3 ^l	D4 ^m	C1 ⁿ	C2 ^o	C3 ^p
$\bar{x}$, mg/100 g	95,9	77,7	66,9	139,2	87,2	75,1	69,3	99,2	99,9	212,4	617,1	6 088,8	1 555,5	468,0	769,7	325,7
<i>r</i> , mg/100 g	1,86	1,41	4,64	3,36	3,46	2,24	5,76	4,71	2,56	5,91	40,86	147,41	57,37	23,03	64,20	20,97
<i>R</i> , mg/100 g	12,44	8,86	9,97	15,72	12,33	6,67	8,05	10,16	9,52	30,49	87,02	960,12	247,49	57,89	89,40	49,11

NOTE 1 Samples S5, D1, D2, D3, D4, C1, C2 and C3 are expressed as mg/100 g “as is”. Samples S1, S2, S3, S4, S6, S7, S8 and S9 are expressed as mg/100 g reconstituted final product.

NOTE 2 SRM 1869, 1549a and 3233 are trade names of products supplied by the National Institute of Standards and Technology (NIST), US Department of Commerce. This information is given for the convenience of users of this document and does not constitute an endorsement by ISO or IDF of the product named. Equivalent products may be used if they can be shown to lead to the same results.

NOTE 3 M-0142, M0-0614 and CA-0904 are products supplied by Muva Kempton, GmbH. This information is given for the convenience of users of this document and does not constitute an endorsement by ISO or IDF of the product named. Equivalent products may be used if they can be shown to lead to the same results.

^a SRM 1869 (infant/adult nutritional formula II); ^b Infant formula (IF), partially hydrolysed, milk-based; ^c IF, partially hydrolysed, soy-based; ^d IF, elemental, amino-acid-based; ^e IF, ready-to-feed, milk-based; ^f IF, milk-based; ^g IF, soy-based; ^h Toddler formula, milk-based; ⁱ Adult nutritional powder, low fat; ^j M-0142 (UHT skimmed milk); ^k MO-0614 (whey powder); ^l CA-0904 (sodium caseinate); ^m SRM 1549a (whole milk powder); ⁿ Bran pet food; ^o Dry pet food; ^p SRM 3233 (fortified breakfast cereal).

NOTE 1 Samples S5, D1, D2, D3, D4, C1, C2 and C3 are expressed as mg/100 g "as is". Samples S1, S2, S3, S4, S6, S7, S8 and S9 are expressed as mg/100 g reconstituted final product.

NOTE 2 SRM 1869, 1549a and 3233 are trade names of products supplied by the National Institute of Standards and Technology (NIST), US Department of Commerce. This information is given for the convenience of users of this document and does not constitute an endorsement by ISO or IDF of the product named. Equivalent products may be used if they can be shown to lead to the same results.

NOTE 3 M-0142, MO-0614 and CA-0904 are products supplied by Muva Kempten, GmbH. This information is given for the convenience of users of this document and does not constitute an endorsement by ISO or IDF of the product named. Equivalent products may be used if they can be shown to lead to the same results.

11 Test report

The test report shall specify:

- a) all the information required for the complete identification of the sample;
- b) the sampling method used, if known;
- c) the method used, with reference to this document, i.e. ISO 4214 | IDF 254;
- d) all operating details not specified in this document, or regarded as optional, together with details of any incident which might have influenced the result(s);
- e) the test result(s) obtained and, if the repeatability or the recovery has been checked, the final quoted result obtained;
- f) the date of the test.

Annex A **(informative)**

Precision data

The data given in [Tables A.1](#) to [A.17](#) were obtained in an international interlaboratory study, and published in 2021^[2] in accordance with ISO 5725-2^[3] and the AOAC-IUPAC Harmonized Protocol for collaborative study procedures,^[4] to assess precision characteristics of a method of analysis. The study was performed based on requirements for collaborative study procedures to validate characteristics of a method of analysis.^[5] Those requirements applied only to samples S1 to S9.

The results of samples S5, D1, D2, D3, D4, C1, C2 and C3 are expressed as mg/100 g sample “as is”. Samples S1, S2, S3, S4, S6, S7, S8 and S9 are expressed as mg/100 g reconstituted final product (25 g powder plus 200 g water).

Table A.1 — Precision data for alanine

Sample	S1 ^a	S2 ^b	S3 ^c	S4 ^d	S5 ^e	S6 ^f	S7 ^g	S8 ^h	S9 ⁱ	D1 ^j	D2 ^k	D3 ^l	D4 ^m	C1 ⁿ	C2 ^o	C3 ^p
Year of interlaboratory test	2020	2020	2020	2020	2020	2020	2020	2020	2020	2020	2020	2020	2020	2020	2020	2020
Number of laboratories retained after eliminating outliers	11	10	10	10	9	11	10	11	11	9	10	10	10	10	10	10
Number of outliers (laboratories)	0	1	1	1	0	0	1	0	0	0	0	0	0	0	0	0
Number of accepted results	11	11	11	11	9	11	11	11	11	9	10	10	10	10	10	10
Mean value, $\bar{x}$, mg/100 g	60,0	66,5	59,9	60,9	59,7	49,4	60,9	59,7	86,7	107,8	516,1	2 756,4	816,7	421,5	756,4	271,5
Repeatability standard deviation, s_r , mg/100 g	1,41	0,79	1,04	0,86	0,96	0,52	0,92	1,14	1,28	2,68	13,58	38,63	16,56	4,77	19,24	5,59
Reproducibility standard deviation, s_R , mg/100 g	5,33	3,03	2,71	3,32	2,49	2,07	3,03	2,67	4,84	5,80	27,63	195,17	44,02	18,57	40,64	10,83
Coefficient of variation of repeatability, $C_{V,r}$, %	2,35	1,18	1,73	1,41	1,61	1,06	1,51	1,91	1,48	2,49	2,63	1,40	2,03	1,13	2,54	2,06
Coefficient of variation of reproducibility, $C_{V,R}$, %	8,89	4,55	4,53	5,46	4,16	4,18	4,98	4,47	5,58	5,38	5,35	7,08	5,39	4,41	5,37	3,99
Repeatability limit, r [$r = 2,8 \times s_r$], mg/100 g	3,94	2,21	2,91	2,41	2,70	1,46	2,57	3,20	3,59	7,52	38,03	108,16	46,37	13,37	53,87	15,64
Reproducibility limit, R [$R = 2,8 \times s_R$], mg/100 g	14,93	8,47	7,60	9,31	6,96	5,79	8,48	7,47	13,56	16,23	77,36	546,48	123,27	52,01	113,78	30,31
Recovery, %	100												97			86
HorRat value	1,46	0,76	0,74	0,90	0,68	0,67	0,82	0,73	0,97	0,96	1,21	2,06	1,31	0,97	1,29	0,82

^a SRM 1869 (infant/adult nutritional formula II); ^b IF, partially hydrolysed, milk-based; ^c IF, partially hydrolysed, soy-based; ^d IF, elemental, amino-acid-based; ^e IF, ready-to-feed, milk-based; ^f IF, milk-based; ^g IF, soy-based; ^h Toddler formula, milk-based; ⁱ Adult nutritional powder, low fat; ^j M-0142 (UHT skimmed milk); ^k MO-0614 (whey powder); ^l CA-0904 (sodium caseinate); ^m SRM 1549a (whole milk powder); ⁿ Bran pet food; ^o Dry pet food; ^p SRM 3233 (fortified breakfast cereal).

NOTE 1 Samples S5, D1, D2, D3, D4, C1, C2 and C3 are expressed as mg/100 g "as is". Samples S1, S2, S3, S4, S6, S7, S8 and S9 are expressed as mg/100 g reconstituted final product.

NOTE 2 SRM 1869, 1549a and 3233 are trade names of products supplied by NIST. This information is given for the convenience of users of this document and does not constitute an endorsement by ISO or IDF of the product named. Equivalent products may be used if they can be shown to lead to the same results.

NOTE 3 M-0142, MO-0614 and CA-0904 are products supplied by Muva Kempten GmbH. This information is given for the convenience of users of this document and does not constitute an endorsement by ISO or IDF of the product named. Equivalent products may be used if they can be shown to lead to the same results.

Table A.2 — Precision data for arginine

Sample	S1 ^a	S2 ^b	S3 ^c	S4 ^d	S5 ^e	S6 ^f	S7 ^g	S8 ^h	S9 ⁱ	D1 ^j	D2 ^k	D3 ^l	D4 ^m	C1 ⁿ	C2 ^o	C3 ^p
Year of interlaboratory test	2020	2020	2020	2020	2020	2020	2020	2020	2020	2020	2020	2020	2020	2020	2020	2020
Number of laboratories retained after eliminating outliers	11	10	10	10	7	11	11	11	9	9	9	10	10	8	10	9
Number of outliers (laboratories)	0	1	1	1	2	0	0	0	2	0	1	0	0	2	0	1
Number of accepted results	11	11	11	11	9	11	11	11	11	9	10	10	10	10	10	10
Mean value, $\bar{x}$, mg/100 g	62,8	32,7	102,3	106,4	32,9	36,5	103,5	50,6	89,1	115,0	256,4	3 448,9	850,7	463,3	839,1	254,2
Repeatability standard deviation, s_r , mg/100 g	2,30	0,87	1,88	2,14	1,47	1,12	2,18	0,89	2,42	4,33	18,88	61,37	18,23	9,72	37,47	6,98
Reproducibility standard deviation, s_R , mg/100 g	5,16	1,54	4,93	4,07	1,84	2,08	4,74	2,10	3,59	9,81	18,88	202,11	32,56	32,82	48,41	20,83
Coefficient of variation of repeatability, $C_{V,r}$, %	3,66	2,65	1,84	2,01	4,47	3,07	2,11	1,76	2,71	3,76	7,36	1,78	2,14	2,10	4,47	2,74
Coefficient of variation of reproducibility, $C_{V,R}$, %	8,22	4,72	4,81	3,82	5,61	5,69	4,58	4,16	4,03	8,53	7,36	5,86	3,83	7,08	5,77	8,19
Repeatability limit, r [$r = 2,8 \times s_r$], mg/100 g	6,44	2,42	5,26	5,99	4,12	3,14	6,12	2,48	6,76	12,12	52,87	171,83	51,03	27,21	104,91	19,54
Reproducibility limit, R [$R = 2,8 \times s_R$], mg/100 g	14,45	4,32	13,80	11,39	5,16	5,82	13,28	5,89	10,06	27,48	52,87	565,90	91,17	91,90	135,54	58,31
Recovery, %	99												96			80
HorRat value	1,35	0,71	0,85	0,68	0,84	0,86	0,81	0,66	0,70	1,54	1,50	1,77	0,93	1,58	1,40	1,67

^a SRM 1869 (infant/adult nutritional formula II); ^b IF, partially hydrolysed, milk-based; ^c IF, partially hydrolysed, soy-based; ^d IF, elemental, amino-acid-based; ^e IF, ready-to-feed, milk-based; ^f IF, milk-based; ^g IF, soy-based; ^h Toddler formula, milk-based; ⁱ Adult nutritional powder, low fat; ^j M-0142 (UHT skimmed milk); ^k MO-0614 (whey powder); ^l CA-0904 (sodium caseinate); ^m SRM 1549a (whole milk powder); ⁿ Bran pet food; ^o Dry pet food; ^p SRM 3233 (fortified breakfast cereal).

NOTE 1 Samples S5, D1, D2, D3, D4, C1, C2 and C3 are expressed as mg/100 g "as is". Samples S1, S2, S3, S4, S6, S7, S8 and S9 are expressed as mg/100 g reconstituted final product.

NOTE 2 SRM 1869, 1549a and 3233 are trade names of products supplied by NIST. This information is given for the convenience of users of this document and does not constitute an endorsement by ISO or IDF of the product named. Equivalent products may be used if they can be shown to lead to the same results.

NOTE 3 M-0142, MO-0614 and CA-0904 are products supplied by Muva Kempton GmbH. This information is given for the convenience of users of this document and does not constitute an endorsement by ISO or IDF of the product named. Equivalent products may be used if they can be shown to lead to the same results.

Table A.3 — Precision data for asparagine and aspartic acid (combined)

Sample	S1 ^a	S2 ^b	S3 ^c	S4 ^d	S5 ^e	S6 ^f	S7 ^g	S8 ^h	S9 ⁱ	D1 ^j	D2 ^k	D3 ^l	D4 ^m	C1 ⁿ	C2 ^o	C3 ^p
Year of interlaboratory test	2020	2020	2020	2020	2020	2020	2020	2020	2020	2020	2020	2020	2020	2020	2020	2020
Number of laboratories retained after eliminating outliers	9	11	10	11	9	11	11	11	10	9	9	10	10	10	10	9
Number of outliers (laboratories)	2	0	1	0	0	0	0	0	1	0	1	0	0	0	0	1
Number of accepted results	11	11	11	11	9	11	11	11	11	9	10	10	10	10	10	10
Mean value, $\bar{x}$, mg/100 g	143,1	150,7	163,8	182,0	136,9	112,8	166,8	139,0	137,2	259,3	1 148,0	6 380,1	1 950,0	619,6	1 791,7	374,4
Repeatability standard deviation, s_r , mg/100 g	1,24	4,24	4,09	9,01	3,11	1,42	4,57	2,62	2,45	7,02	24,04	115,91	61,82	9,06	54,62	10,02
Reproducibility standard deviation, s_R , mg/100 g	6,95	6,29	6,44	9,24	9,49	4,69	7,81	5,29	7,59	23,18	55,50	550,09	90,37	30,98	132,46	21,60
Coefficient of variation of repeatability, $C_{V,r}$, %	0,86	2,81	2,50	4,95	2,27	1,26	2,74	1,88	1,79	2,71	2,09	1,82	3,17	1,46	3,05	2,68
Coefficient of variation of reproducibility, $C_{V,R}$, %	4,86	4,17	3,93	5,08	6,94	4,16	4,68	3,80	5,53	8,94	4,83	8,62	4,63	5,00	7,39	5,77
Repeatability limit, r [$r = 2,8 \times s_r$], mg/100 g	3,47	11,87	11,46	25,22	8,70	3,98	12,80	7,34	6,86	19,66	67,31	324,55	173,10	25,38	152,94	28,05
Reproducibility limit, R [$R = 2,8 \times s_R$], mg/100 g	19,46	17,61	18,03	25,86	26,58	13,14	21,86	14,80	21,25	64,90	155,39	1 540,26	253,03	86,73	370,90	60,48
Recovery, %	99												99			87
HorRat value	0,91	0,78	0,75	0,98	1,29	0,75	0,89	0,71	1,03	1,82	1,23	2,85	1,28	1,16	2,02	1,24

^a SRM 1869 (infant/adult nutritional formula II); ^b IF, partially hydrolysed, milk-based; ^c IF, partially hydrolysed, soy-based; ^d IF, elemental, amino-acid-based; ^e IF, ready-to-feed, milk-based; ^f IF, milk-based; ^g IF, soy-based; ^h Toddler formula, milk-based; ⁱ Adult nutritional powder, low fat; ^j M-0142 (UHT skimmed milk); ^k MO-0614 (whey powder); ^l CA-0904 (sodium caseinate); ^m SRM 1549a (whole milk powder); ⁿ Bran pet food; ^o Dry pet food; ^p SRM 3233 (fortified breakfast cereal).

NOTE 1 Samples S5, D1, D2, D3, D4, C1, C2 and C3 are expressed as mg/100 g "as is". Samples S1, S2, S3, S4, S6, S7, S8 and S9 are expressed as mg/100 g reconstituted final product.

NOTE 2 SRM 1869, 1549a and 3233 are trade names of products supplied by NIST. This information is given for the convenience of users of this document and does not constitute an endorsement by ISO or IDF of the product named. Equivalent products may be used if they can be shown to lead to the same results.

NOTE 3 M-0142, MO-0614 and CA-0904 are products supplied by Muva Kempten GmbH. This information is given for the convenience of users of this document and does not constitute an endorsement by ISO or IDF of the product named. Equivalent products may be used if they can be shown to lead to the same results.

Table A.4 — Precision data for cysteine and cystine (combined)

Sample	S1 ^a	S2 ^b	S3 ^c	S4 ^d	S5 ^e	S6 ^f	S7 ^g	S8 ^h	S9 ⁱ	D1 ^j	D2 ^k	D3 ^l	D4 ^m	C1 ⁿ	C2 ^o	C3 ^p
Year of interlaboratory test	2020	2020	2020	2020	2020	2020	2020	2020	2020	2020	2020	2020	2020	2020	2020	2020
Number of laboratories retained after eliminating outliers	9	8	8	9	8	9	9	8	9	8	7	9	8	9	9	9
Number of outliers (laboratories)	0	1	1	0	0	0	0	1	1	0	2	0	1	0	0	0
Number of accepted results	9	9	9	9	8	9	9	9	10	8	9	9	9	9	9	9
Mean value, $\bar{x}$, mg/100 g	18,1	35,1	18,0	24,4	20,3	19,6	18,0	20,9	16,7	25,9	258,9	389,1	199,8	177,2	187,8	125,6
Repeatability standard deviation, s_r , mg/100 g	0,98	0,59	0,49	1,41	0,97	0,42	0,82	0,26	0,30	0,73	4,43	9,97	7,01	5,61	7,36	2,89
Reproducibility standard deviation, s_R , mg/100 g	2,48	4,00	1,99	2,92	2,04	2,31	2,23	1,36	2,59	1,83	17,01	42,39	7,68	17,38	10,95	10,83
Coefficient of variation of repeatability, $C_{V,r}$, %	5,38	1,69	2,72	5,77	4,75	2,16	4,58	1,27	1,79	2,84	1,71	2,56	3,51	3,16	3,92	2,30
Coefficient of variation of reproducibility, $C_{V,R}$, %	13,71	11,39	11,08	11,94	10,01	11,75	12,38	6,49	15,51	7,08	6,57	10,89	3,84	9,81	5,83	8,62
Repeatability limit, r [$r = 2,8 \times s_r$], mg/100 g	2,73	1,67	1,37	3,95	2,71	1,19	2,31	0,74	0,84	2,06	12,40	27,93	19,63	15,70	20,60	8,10
Reproducibility limit, R [$R = 2,8 \times s_R$], mg/100 g	6,95	11,20	5,57	8,17	5,70	6,46	6,24	3,80	7,24	5,13	47,63	118,68	21,50	48,67	30,66	30,33
Recovery, %	109												111			83
HorRat value	1,87	1,72	1,51	1,71	1,39	1,63	1,69	0,91	2,09	1,02	1,34	2,36	0,75	1,89	1,13	1,58

^a SRM 1869 (infant/adult nutritional formula II); ^b IF, partially hydrolysed, milk-based; ^c IF, partially hydrolysed, soy-based; ^d IF, elemental, amino-acid-based; ^e IF, ready-to-feed, milk-based; ^f IF, milk-based; ^g IF, soy-based; ^h Toddler formula, milk-based; ⁱ Adult nutritional powder, low fat; ^j M-0142 (UHT skimmed milk); ^k MO-0614 (whey powder); ^l CA-0904 (sodium caseinate); ^m SRM 1549a (whole milk powder); ⁿ Bran pet food; ^o Dry pet food; ^p SRM 3233 (fortified breakfast cereal).

NOTE 1 Samples S5, D1, D2, D3, D4, C1, C2 and C3 are expressed as mg/100 g “as is”. Samples S1, S2, S3, S4, S6, S7, S8 and S9 are expressed as mg/100 g reconstituted final product.

NOTE 2 SRM 1869, 1549a and 3233 are trade names of products supplied by NIST. This information is given for the convenience of users of this document and does not constitute an endorsement by ISO or IDF of the product named. Equivalent products may be used if they can be shown to lead to the same results.

NOTE 3 M-0142, MO-0614 and CA-0904 are products supplied by Muva Kempton GmbH. This information is given for the convenience of users of this document and does not constitute an endorsement by ISO or IDF of the product named. Equivalent products may be used if they can be shown to lead to the same results.

Table A.5 — Precision data for glutamine and glutamic acid (combined)

Sample	S1 ^a	S2 ^b	S3 ^c	S4 ^d	S5 ^e	S6 ^f	S7 ^g	S8 ^h	S9 ⁱ	D1 ^j	D2 ^k	D3 ^l	D4 ^m	C1 ⁿ	C2 ^o	C3 ^p
Year of interlaboratory test	2020	2020	2020	2020	2020	2020	2020	2020	2020	2020	2020	2020	2020	2020	2020	2020
Number of laboratories retained after eliminating outliers	9	10	10	9	9	11	10	11	11	9	10	10	10	10	10	9
Number of outliers (laboratories)	2	1	1	2	0	0	1	0	0	0	0	0	0	0	0	1
Number of accepted results	11	11	11	11	9	11	11	11	11	9	10	10	10	10	10	10
Mean value, $\bar{x}$, mg/100 g	336,7	242,4	276,8	234,7	279,3	248,4	282,9	334,3	242,7	720,0	1 949,6	20 287,3	5 389,8	2 663,7	2 313,9	2 010,7
Repeatability standard deviation, s_r , mg/100 g	2,52	3,46	5,37	2,24	4,24	2,21	3,50	6,36	3,87	17,34	66,75	332,32	111,38	33,54	86,34	43,40
Reproducibility standard deviation, s_R , mg/100 g	12,05	8,72	12,15	8,53	9,93	6,74	8,32	10,81	9,58	48,59	80,68	1 848,19	244,60	132,45	134,37	103,71
Coefficient of variation of repeatability, $C_{V,r}$, %	0,75	1,43	1,94	0,95	1,52	0,89	1,24	1,90	1,59	2,41	3,42	1,64	2,07	1,26	3,73	2,16
Coefficient of variation of reproducibility, $C_{V,R}$, %	3,58	3,60	4,39	3,63	3,56	2,71	2,94	3,23	3,95	6,75	4,14	9,11	4,54	4,97	5,81	5,16
Repeatability limit, r [$r = 2,8 \times s_r$], mg/100 g	7,05	9,68	15,03	6,27	11,88	6,18	9,80	17,82	10,84	48,54	186,89	930,50	311,87	93,90	241,76	121,51
Reproducibility limit, R [$R = 2,8 \times s_R$], mg/100 g	33,74	24,42	34,03	23,87	27,81	18,87	23,30	30,25	26,81	136,04	225,89	5 174,93	684,89	370,85	376,24	290,40
Recovery, %	103												101			91
HorRat value	0,76	0,73	0,90	0,73	0,73	0,55	0,61	0,69	0,80	1,61	1,14	3,58	1,46	1,44	1,65	1,43

^a SRM 1869 (infant/adult nutritional formula II); ^b IF, partially hydrolysed, milk-based; ^c IF, partially hydrolysed, soy-based; ^d IF, elemental, amino-acid-based; ^e IF, ready-to-feed, milk-based; ^f IF, milk-based; ^g IF, soy-based; ^h Toddler formula; ⁱ Adult nutritional powder, low fat; ^j M-0142 (UHT skimmed milk); ^k MO-0614 (whey powder); ^l CA-0904 (sodium caseinate); ^m SRM 1549a (whole milk powder); ⁿ Bran pet food; ^o Dry pet food; ^p SRM 3233 (fortified breakfast cereal).

NOTE 1 Samples S5, D1, D2, D3, D4, C1, C2 and C3 are expressed as mg/100 g "as is". Samples S1, S2, S3, S4, S6, S7, S8 and S9 are expressed as mg/100 g reconstituted final product.

NOTE 2 SRM 1869, 1549a and 3233 are trade names of products supplied by NIST. This information is given for the convenience of users of this document and does not constitute an endorsement by ISO or IDF of the product named. Equivalent products may be used if they can be shown to lead to the same results.

NOTE 3 M-0142, MO-0614 and CA-0904 are products supplied by Muva Kempten GmbH. This information is given for the convenience of users of this document and does not constitute an endorsement by ISO or IDF of the product named. Equivalent products may be used if they can be shown to lead to the same results.

Table A.6 — Precision data for glycine

Sample	S1 ^a	S2 ^b	S3 ^c	S4 ^d	S5 ^e	S6 ^f	S7 ^g	S8 ^h	S9 ⁱ	D1 ^j	D2 ^k	D3 ^l	D4 ^m	C1 ⁿ	C2 ^o	C3 ^p
Year of interlaboratory test	2020	2020	2020	2020	2020	2020	2020	2020	2020	2020	2020	2020	2020	2020	2020	2020
Number of laboratories retained after eliminating outliers	9	10	10	11	7	11	11	11	9	8	10	10	10	10	9	10
Number of outliers (laboratories)	2	1	1	0	2	0	0	0	2	1	0	0	0	0	1	0
Number of accepted results	11	11	11	11	9	11	11	11	11	9	10	10	10	10	10	10
Mean value, $\bar{x}$, mg/100 g	36,2	24,7	57,9	46,8	27,4	22,6	58,3	30,6	150,7	62,2	228,0	1 745,0	471,3	495,7	789,3	313,0
Repeatability standard deviation, s_r , mg/100 g	0,58	0,31	1,23	1,52	0,65	0,35	1,34	0,68	2,24	2,68	13,09	43,78	14,49	7,63	7,95	7,94
Reproducibility standard deviation, s_R , mg/100 g	2,97	1,48	2,57	2,68	1,12	1,24	3,23	1,71	9,03	3,27	17,33	116,89	21,19	30,32	23,72	17,32
Coefficient of variation of repeatability, $C_{V,r}$, %	1,59	1,27	2,12	3,26	2,36	1,54	2,30	2,22	1,49	4,30	5,74	2,51	3,07	1,54	1,01	2,54
Coefficient of variation of reproducibility, $C_{V,R}$, %	8,19	5,99	4,44	5,73	4,09	5,46	5,54	5,58	5,99	5,25	7,60	6,70	4,50	6,12	3,01	5,53
Repeatability limit, r [$r = 2,8 \times s_r$], mg/100 g	1,61	0,88	3,43	4,27	1,81	0,97	3,76	1,91	6,28	7,50	36,65	122,58	40,57	21,36	22,27	22,24
Reproducibility limit, R [$R = 2,8 \times s_R$], mg/100 g	8,30	4,15	7,19	7,51	3,14	3,46	9,04	4,79	25,29	9,16	48,52	327,30	59,32	84,88	66,42	48,49
Recovery, %	100												102			93
HorRat value	1,24	0,86	0,72	0,90	0,60	0,77	0,90	0,83	1,13	0,86	1,52	1,82	1,00	1,38	0,73	1,16

^a SRM 1869 (infant/adult nutritional formula II); ^b IF, partially hydrolysed, milk-based; ^c IF, partially hydrolysed, soy-based; ^d IF, elemental, amino-acid-based; ^e IF, ready-to-feed, milk-based; ^f IF, milk-based; ^g IF, soy-based; ^h Toddler formula, milk-based; ⁱ Adult nutritional powder, low fat; ^j M-0142 (UHT skimmed milk); ^k MO-0614 (whey powder); ^l CA-0904 (sodium caseinate); ^m SRM 1549a (whole milk powder); ⁿ Bran pet food; ^o Dry pet food; ^p SRM 3233 (fortified breakfast cereal).

NOTE 1 Samples S5, D1, D2, D3, D4, C1, C2 and C3 are expressed as mg/100 g “as is”. Samples S1, S2, S3, S4, S6, S7, S8 and S9 are expressed as mg/100 g reconstituted final product.

NOTE 2 SRM 1869, 1549a and 3233 are trade names of products supplied by NIST. This information is given for the convenience of users of this document and does not constitute an endorsement by ISO or IDF of the product named. Equivalent products may be used if they can be shown to lead to the same results.

NOTE 3 M-0142, MO-0614 and CA-0904 are products supplied by Muva Kempton GmbH. This information is given for the convenience of users of this document and does not constitute an endorsement by ISO or IDF of the product named. Equivalent products may be used if they can be shown to lead to the same results.

Table A.7 — Precision data for histidine

Sample	S1 ^a	S2 ^b	S3 ^c	S4 ^d	S5 ^e	S6 ^f	S7 ^g	S8 ^h	S9 ⁱ	D1 ^j	D2 ^k	D3 ^l	D4 ^m	C1 ⁿ	C2 ^o	C3 ^p
Year of interlaboratory test	2020	2020	2020	2020	2020	2020	2020	2020	2020	2020	2020	2020	2020	2020	2020	2020
Number of laboratories retained after eliminating outliers	9	9	11	9	7	11	10	9	10	7	10	10	10	10	9	9
Number of outliers (laboratories)	2	2	0	2	2	0	1	2	1	2	0	0	0	0	1	1
Number of accepted results	11	11	11	11	9	11	11	11	11	9	10	10	10	10	10	10
Mean value, $\bar{x}$, mg/100 g	40,3	24,0	35,1	44,0	24,8	27,2	34,8	40,2	43,9	88,5	187,2	2706,5	668,5	241,8	365,3	147,8
Repeatability standard deviation, s_r , mg/100 g	0,69	0,17	2,12	0,88	0,49	0,46	1,08	0,34	0,58	1,18	6,10	82,23	13,39	6,05	8,21	5,13
Reproducibility standard deviation, s_R , mg/100 g	3,71	1,59	2,97	2,14	0,99	2,24	1,90	2,76	1,88	3,37	11,78	214,34	26,47	16,77	14,47	10,63
Coefficient of variation of repeatability, $C_{V,r}$, %	1,71	0,69	6,04	1,99	1,96	1,69	3,09	0,84	1,32	1,33	3,26	3,04	2,00	2,50	2,25	3,47
Coefficient of variation of reproducibility, $C_{V,R}$, %	9,19	6,62	8,48	4,86	3,97	8,25	5,45	6,85	4,28	3,80	6,29	7,92	3,96	6,94	3,96	7,19
Repeatability limit, r [$r = 2,8 \times s_r$], mg/100 g	1,93	0,47	5,93	2,45	1,36	1,29	3,01	0,95	1,62	3,30	17,08	230,24	37,49	16,95	22,98	14,35
Reproducibility limit, R [$R = 2,8 \times s_R$], mg/100 g	10,38	4,46	8,32	5,99	2,76	6,28	5,31	7,72	5,26	9,42	32,98	600,16	74,10	46,97	40,51	29,75
Recovery, %	99												108			93
HorRat value	1,42	0,94	1,28	0,76	0,57	1,20	0,82	1,06	0,67	0,66	1,22	2,30	0,93	1,40	0,85	1,35

^a SRM 1869 (infant/adult nutritional formula 11); ^b IF, partially hydrolysed, milk-based; ^c IF, partially hydrolysed, soy-based; ^d IF, elemental, amino-acid-based; ^e IF, ready-to-feed, milk-based; ^f IF, milk-based; ^g IF, soy-based; ^h Toddler formula, milk-based; ⁱ Adult nutritional powder, low fat; ^j M-0142 (UHT skimmed milk); ^k MO-0614 (whey powder); ^l CA-0904 (sodium caseinate); ^m SRM 1549a (whole milk powder); ⁿ Bran pet food; ^o Dry pet food; ^p SRM 3233 (fortified breakfast cereal).

NOTE 1 Samples S5, D1, D2, D3, D4, C1, C2 and C3 are expressed as mg/100 g "as is". Samples S1, S2, S3, S4, S6, S7, S8 and S9 are expressed as mg/100 g reconstituted final product.

NOTE 2 SRM 1869, 1549a and 3233 are trade names of products supplied by NIST. This information is given for the convenience of users of this document and does not constitute an endorsement by ISO or IDF of the product named. Equivalent products may be used if they can be shown to lead to the same results.

NOTE 3 M-0142, MO-0614 and CA-0904 are products supplied by Muva Kempten GmbH. This information is given for the convenience of users of this document and does not constitute an endorsement by ISO or IDF of the product named. Equivalent products may be used if they can be shown to lead to the same results.

Table A.8 — Precision data for isoleucine

Sample	S1 ^a	S2 ^b	S3 ^c	S4 ^d	S5 ^e	S6 ^f	S7 ^g	S8 ^h	S9 ⁱ	D1 ^j	D2 ^k	D3 ^l	D4 ^m	C1 ⁿ	C2 ^o	C3 ^p
Year of interlaboratory test	2020	2020	2020	2020	2020	2020	2020	2020	2020	2020	2020	2020	2020	2020	2020	2020
Number of laboratories retained after eliminating outliers	10	10	10	10	9	11	11	11	10	9	8	10	10	10	10	10
Number of outliers (laboratories)	1	1	1	1	0	0	0	0	1	0	2	0	0	0	0	0
Number of accepted results	11	11	11	11	9	11	11	11	11	9	10	10	10	10	10	10
Mean value, $\bar{x}$, mg/100 g	86,0	84,0	66,8	120,9	83,3	69,6	68,1	88,5	87,1	175,9	670,6	4 854,9	1 296,0	352,4	657,6	254,1
Repeatability standard deviation, s_r , mg/100 g	0,88	1,44	1,68	1,48	1,17	1,01	1,92	1,65	0,60	3,85	18,91	60,69	18,71	7,24	20,67	7,78
Reproducibility standard deviation, s_R , mg/100 g	3,37	3,51	3,13	4,14	4,51	2,73	2,50	3,20	2,50	12,89	18,91	276,04	71,31	22,66	34,42	19,96
Coefficient of variation of repeatability, $C_{V,r}$, %	1,02	1,72	2,52	1,23	1,40	1,45	2,81	1,86	0,69	2,19	2,82	1,25	1,44	2,06	3,14	3,06
Coefficient of variation of reproducibility, $C_{V,R}$, %	3,92	4,18	4,68	3,42	5,42	3,92	3,67	3,62	2,88	7,33	2,82	5,69	5,50	6,43	5,23	7,86
Repeatability limit, r [$r = 2,8 \times s_r$], mg/100 g	2,47	4,04	4,70	4,15	3,27	2,83	5,36	4,61	1,69	10,78	52,94	169,94	52,39	20,28	57,87	21,80
Reproducibility limit, R [$R = 2,8 \times s_R$], mg/100 g	9,45	9,83	8,75	11,59	12,64	7,65	7,00	8,96	7,01	36,10	52,94	772,90	199,66	63,44	96,38	55,89
Recovery, %	100												116			96
HorRat value	0,68	0,72	0,78	0,62	0,93	0,66	0,61	0,63	0,50	1,41	0,66	1,80	1,43	1,37	1,23	1,60

^a SRM 1869 (infant/adult nutritional formula II); ^b IF, partially hydrolysed, milk-based; ^c IF, partially hydrolysed, soy-based; ^d IF, elemental, amino-acid-based; ^e IF, ready-to-feed, milk-based; ^f IF, milk-based; ^g IF, soy-based; ^h Toddler formula, milk-based; ⁱ Adult nutritional powder, low fat; ^j M-0142 (UHT skimmed milk); ^k MO-0614 (whey powder); ^l CA-0904 (sodium caseinate); ^m SRM 1549a (whole milk powder); ⁿ Bran pet food; ^o Dry pet food; ^p SRM 3233 (fortified breakfast cereal).

NOTE 1 Samples S5, D1, D2, D3, D4, C1, C2 and C3 are expressed as mg/100 g “as is”. Samples S1, S2, S3, S4, S6, S7, S8 and S9 are expressed as mg/100 g reconstituted final product.

NOTE 2 SRM 1869, 1549a and 3233 are trade names of products supplied by NIST. This information is given for the convenience of users of this document and does not constitute an endorsement by ISO or IDF of the product named. Equivalent products may be used if they can be shown to lead to the same results.

NOTE 3 M-0142, MO-0614 and CA-0904 are products supplied by Muva Kempton GmbH. This information is given for the convenience of users of this document and does not constitute an endorsement by ISO or IDF of the product named. Equivalent products may be used if they can be shown to lead to the same results.

Table A.9 — Precision data for leucine

Sample	S1 ^a	S2 ^b	S3 ^c	S4 ^d	S5 ^e	S6 ^f	S7 ^g	S8 ^h	S9 ⁱ	D1 ^j	D2 ^k	D3 ^l	D4 ^m	C1 ⁿ	C2 ^o	C3 ^p
Year of interlaboratory test	2020	2020	2020	2020	2020	2020	2020	2020	2020	2020	2020	2020	2020	2020	2020	2020
Number of laboratories retained after eliminating outliers	10	10	9	10	8	11	11	11	10	8	10	10	10	10	10	9
Number of outliers (laboratories)	1	1	2	1	1	0	0	0	1	1	0	0	0	0	0	1
Number of accepted results	11	11	11	11	9	11	11	11	11	9	10	10	10	10	10	10
Mean value, $\bar{x}$, mg/100 g	153,0	141,3	110,9	197,0	147,4	124,7	113,4	161,6	138,0	331,3	1 104,1	8 660,8	2 424,4	672,9	1 151,8	484,8
Repeatability standard deviation, s_r , mg/100 g	1,82	1,33	2,24	2,40	0,78	1,63	3,39	2,74	1,29	2,81	30,43	102,42	36,79	4,99	30,25	9,49
Reproducibility standard deviation, s_R , mg/100 g	5,20	4,34	2,93	6,49	6,02	3,91	3,78	4,33	3,32	11,41	35,84	363,08	85,69	25,89	36,65	21,77
Coefficient of variation of repeatability, $C_{V,r}$, %	1,19	0,94	2,01	1,22	0,53	1,31	2,99	1,69	0,93	0,85	2,76	1,18	1,52	0,74	2,63	1,96
Coefficient of variation of reproducibility, $C_{V,R}$, %	3,40	3,07	2,64	3,29	4,09	3,14	3,33	2,68	2,40	3,44	3,25	4,19	3,53	3,85	3,18	4,49
Repeatability limit, r [$r = 2,8 \times s_r$], mg/100 g	5,09	3,74	6,26	6,71	2,18	4,58	9,50	7,66	3,61	7,87	85,20	286,78	103,00	13,98	84,69	26,57
Reproducibility limit, R [$R = 2,8 \times s_R$], mg/100 g	14,57	12,14	8,21	18,16	16,86	10,96	10,58	12,12	9,29	31,93	100,36	1 016,61	239,93	72,48	102,61	60,96
Recovery, %	99												101			90
HorRat value	0,64	0,57	0,47	0,64	0,77	0,57	0,60	0,51	0,45	0,73	0,82	1,45	1,01	0,91	0,81	1,01

^a SRM 1869 (infant/adult nutritional formula II); ^b IF, partially hydrolysed, milk-based; ^c IF, partially hydrolysed, soy-based; ^d IF, elemental, amino-acid-based; ^e IF, ready-to-feed, milk-based; ^f IF, milk-based; ^g IF, soy-based; ^h Toddler formula, milk-based; ⁱ Adult nutritional powder, low fat; ^j M-0142 (UHT skimmed milk); ^k MO-0614 (whey powder); ^l CA-0904 (sodium caseinate); ^m SRM 1549a (whole milk powder); ⁿ Bran pet food; ^o Dry pet food; ^p SRM 3233 (fortified breakfast cereal).

NOTE 1 Samples S5, D1, D2, D3, D4, C1, C2 and C3 are expressed as mg/100 g "as is". Samples S1, S2, S3, S4, S6, S7, S8 and S9 are expressed as mg/100 g reconstituted final product.

NOTE 2 SRM 1869, 1549a and 3233 are trade names of products supplied by NIST. This information is given for the convenience of users of this document and does not constitute an endorsement by ISO or IDF of the product named. Equivalent products may be used if they can be shown to lead to the same results.

NOTE 3 M-0142, MO-0614 and CA-0904 are products supplied by Muva Kempten GmbH. This information is given for the convenience of users of this document and does not constitute an endorsement by ISO or IDF of the product named. Equivalent products may be used if they can be shown to lead to the same results.

Table A.10 — Precision data for lysine

Sample	S1 ^a	S2 ^b	S3 ^c	S4 ^d	S5 ^e	S6 ^f	S7 ^g	S8 ^h	S9 ⁱ	D1 ^j	D2 ^k	D3 ^l	D4 ^m	C1 ⁿ	C2 ^o	C3 ^p
Year of interlaboratory test	2020	2020	2020	2020	2020	2020	2020	2020	2020	2020	2020	2020	2020	2020	2020	2020
Number of laboratories retained after eliminating outliers	9	11	10	11	8	10	11	11	10	9	10	10	10	10	10	9
Number of outliers (laboratories)	2	0	1	0	1	1	0	0	1	0	0	0	0	0	0	1
Number of accepted results	11	11	11	11	9	11	11	11	11	9	10	10	10	10	10	10
Mean value, $\bar{x}$, mg/100 g	131,4	122,9	86,1	110,2	97,4	99,5	84,2	130,7	91,1	260,1	959,1	7 210,1	1 953,0	190,9	813,2	99,5
Repeatability standard deviation, s_r , mg/100 g	2,70	3,04	2,12	4,84	1,64	1,39	2,32	4,08	1,88	12,14	39,87	188,35	54,39	6,04	29,69	5,43
Reproducibility standard deviation, s_R , mg/100 g	10,86	4,50	3,82	5,17	9,18	3,99	3,65	5,61	4,82	22,89	42,58	564,60	104,64	14,07	35,04	9,55
Coefficient of variation of repeatability, $C_{V,r}$, %	2,05	2,47	2,46	4,39	1,68	1,40	2,75	3,12	2,07	4,67	4,16	2,61	2,78	3,16	3,65	5,46
Coefficient of variation of reproducibility, $C_{V,R}$, %	8,26	3,67	4,43	4,69	9,42	4,01	4,34	4,29	5,29	8,80	4,44	7,83	5,36	7,37	4,31	9,59
Repeatability limit, r [$r = 2,8 \times s_r$], mg/100 g	7,56	8,51	5,93	13,54	4,58	3,89	6,49	11,42	5,27	33,99	111,63	527,39	152,28	16,90	83,12	15,21
Reproducibility limit, R [$R = 2,8 \times s_R$], mg/100 g	30,40	12,61	10,70	14,48	25,69	11,16	10,23	15,71	13,50	64,10	119,22	1 580,88	292,99	39,40	98,12	26,74
Recovery, %	100												95			98
HorRat value	1,52	0,67	0,77	0,84	1,66	0,71	0,75	0,79	0,92	1,80	1,10	2,64	1,48	1,44	1,04	1,69

^a SRM 1869 (infant/adult nutritional formula II); ^b IF, partially hydrolysed, milk-based; ^c IF, partially hydrolysed, soy-based; ^d IF, elemental, amino-acid-based; ^e IF, ready-to-feed, milk-based; ^f IF, milk-based; ^g IF, soy-based; ^h Toddler formula, milk-based; ⁱ Adult nutritional powder, low fat; ^j M-0142 (UHT skimmed milk); ^k MO-0614 (whey powder); ^l CA-0904 (sodium caseinate); ^m SRM 1549a (whole milk powder); ⁿ Bran pet food; ^o Dry pet food; ^p SRM 3233 (fortified breakfast cereal).

NOTE 1 Samples S5, D1, D2, D3, D4, C1, C2 and C3 are expressed as mg/100 g “as is”. Samples S1, S2, S3, S4, S6, S7, S8 and S9 are expressed as mg/100 g reconstituted final product.

NOTE 2 SRM 1869, 1549a and 3233 are trade names of products supplied by NIST. This information is given for the convenience of users of this document and does not constitute an endorsement by ISO or IDF of the product named. Equivalent products may be used if they can be shown to lead to the same results.

NOTE 3 M-0142, MO-0614 and CA-0904 are products supplied by Muva Kempton GmbH. This information is given for the convenience of users of this document and does not constitute an endorsement by ISO or IDF of the product named. Equivalent products may be used if they can be shown to lead to the same results.

Table A.11 — Precision data for methionine

Sample	S1 ^a	S2 ^b	S3 ^c	S4 ^d	S5 ^e	S6 ^f	S7 ^g	S8 ^h	S9 ⁱ	D1 ^j	D2 ^k	D3 ^l	D4 ^m	C1 ⁿ	C2 ^o	C3 ^p
Year of interlaboratory test	2020	2020	2020	2020	2020	2020	2020	2020	2020	2020	2020	2020	2020	2020	2020	2020
Number of laboratories retained after eliminating outliers	9	11	10	10	7	11	11	11	10	9	10	10	10	10	8	9
Number of outliers (laboratories)	2	0	1	1	2	0	0	0	1	0	0	0	0	0	2	1
Number of accepted results	11	11	11	11	9	11	11	11	11	9	10	10	10	10	10	10
Mean value, $\bar{x}$, mg/100 g	51,6	27,9	36,6	43,7	30,2	29,0	42,1	39,9	37,3	87,5	214,9	2 655,2	643,3	136,1	138,3	101,9
Repeatability standard deviation, s_r , mg/100 g	0,59	1,13	0,88	1,02	0,99	0,60	0,42	0,70	0,40	4,05	10,38	57,73	18,39	2,87	8,20	3,32
Reproducibility standard deviation, s_R , mg/100 g	3,02	2,03	1,68	1,94	1,77	1,22	1,81	1,96	1,65	7,51	24,80	157,39	41,25	11,42	26,48	7,19
Coefficient of variation of repeatability, $C_{V,r}$, %	1,14	4,07	2,40	2,35	3,29	2,06	0,99	1,76	1,06	4,63	4,83	2,17	2,86	2,11	5,93	3,26
Coefficient of variation of reproducibility, $C_{V,R}$, %	5,86	7,27	4,59	4,45	5,86	4,21	4,31	4,91	4,44	8,59	11,54	5,93	6,41	8,39	19,15	7,05
Repeatability limit, r [$r = 2,8 \times s_r$], mg/100 g	1,65	3,17	2,46	2,87	2,78	1,67	1,17	1,96	1,11	11,35	29,05	161,65	51,50	8,05	22,97	9,30
Reproducibility limit, R [$R = 2,8 \times s_R$], mg/100 g	8,46	5,67	4,71	5,44	4,97	3,42	5,08	5,48	4,63	21,04	69,45	440,70	115,51	31,98	74,14	20,12
Recovery, %	98												95			75
HorRat value	0,94	1,06	0,70	0,69	0,87	0,62	0,67	0,76	0,68	1,49	2,29	1,72	1,50	1,55	3,55	1,25

^a SRM 1869 (infant/adult nutritional formula II); ^b IF, partially hydrolysed, milk-based; ^c IF, partially hydrolysed, soy-based; ^d IF, elemental, amino-acid-based; ^e IF, ready-to-feed, milk-based; ^f IF, milk-based; ^g IF, soy-based; ^h Toddler formula, milk-based; ⁱ Adult nutritional powder, low fat; ^j M-0142 (UHT skimmed milk); ^k MO-0614 (whey powder); ^l CA-0904 (sodium caseinate); ^m SRM 1549a (whole milk powder); ⁿ Bran pet food; ^o Dry pet food; ^p SRM 3233 (fortified breakfast cereal).

NOTE 1 Samples S5, D1, D2, D3, D4, C1, C2 and C3 are expressed as mg/100 g "as is". Samples S1, S2, S3, S4, S6, S7, S8 and S9 are expressed as mg/100 g reconstituted final product.

NOTE 2 SRM 1869, 1549a and 3233 are trade names of products supplied by NIST. This information is given for the convenience of users of this document and does not constitute an endorsement by ISO or IDF of the product named. Equivalent products may be used if they can be shown to lead to the same results.

NOTE 3 M-0142, MO-0614 and CA-0904 are products supplied by Muva Kempten GmbH. This information is given for the convenience of users of this document and does not constitute an endorsement by ISO or IDF of the product named. Equivalent products may be used if they can be shown to lead to the same results.

Table A.12 — Precision data for phenylalanine

Sample	S1 ^a	S2 ^b	S3 ^c	S4 ^d	S5 ^e	S6 ^f	S7 ^g	S8 ^h	S9 ⁱ	D1 ^j	D2 ^k	D3 ^l	D4 ^m	C1 ⁿ	C2 ^o	C3 ^p
Year of interlaboratory test	2020	2020	2020	2020	2020	2020	2020	2020	2020	2020	2020	2020	2020	2020	2020	2020
Number of laboratories retained after eliminating outliers	9	11	9	10	8	11	11	11	10	7	8	10	10	10	9	9
Number of outliers (laboratories)	2	0	2	1	1	0	0	0	1	2	2	0	0	0	1	1
Number of accepted results	11	11	11	11	9	11	11	11	11	9	10	10	10	10	10	10
Mean value, $\bar{x}$, mg/100 g	73,5	42,9	72,9	100,4	57,4	50,5	73,6	71,4	76,5	164,4	341,4	4 856,9	1 212,2	457,3	735,9	323,0
Repeatability standard deviation, s_r , mg/100 g	0,92	1,60	1,05	2,48	1,85	1,50	2,58	1,60	2,24	2,05	4,81	192,23	28,51	14,82	19,44	8,60
Reproducibility standard deviation, s_R , mg/100 g	2,09	2,46	3,08	3,05	3,26	2,15	2,58	3,23	3,29	5,42	19,06	387,82	52,25	26,48	31,90	17,78
Coefficient of variation of repeatability, $C_{V,r}$, %	1,25	3,74	1,44	2,47	3,23	2,97	3,50	2,24	2,93	1,25	1,41	3,96	2,35	3,24	2,64	2,66
Coefficient of variation of reproducibility, $C_{V,R}$, %	2,84	5,74	4,22	3,04	5,67	4,26	3,50	4,53	4,30	3,30	5,58	7,98	4,31	5,79	4,33	5,51
Repeatability limit, r [$r = 2,8 \times s_r$], mg/100 g	2,58	4,48	2,93	6,94	5,19	4,21	7,21	4,48	6,27	5,74	13,47	538,26	79,83	41,50	54,43	24,08
Reproducibility limit, R [$R = 2,8 \times s_R$], mg/100 g	5,84	6,88	8,62	8,54	9,12	6,02	7,21	9,05	9,20	15,19	53,37	1 085,91	146,31	74,15	89,31	49,79
Recovery, %	97												100			88
HorRat value	0,48	0,89	0,71	0,54	0,92	0,68	0,59	0,76	0,73	0,63	1,19	2,53	1,11	1,29	1,03	1,16

^a SRM 1869 (infant/adult nutritional formula II); ^b IF, partially hydrolysed, milk-based; ^c IF, partially hydrolysed, soy-based; ^d IF, elemental, amino-acid-based; ^e IF, ready-to-feed, milk-based; ^f IF, milk-based; ^g IF, soy-based; ^h Toddler formula, milk-based; ⁱ Adult nutritional powder, low fat; ^j M-0142 (UHT skimmed milk); ^k MO-0614 (whey powder); ^l CA-0904 (sodium caseinate); ^m SRM 1549a (whole milk powder); ⁿ Bran pet food; ^o Dry pet food; ^p SRM 3233 (fortified breakfast cereal).

NOTE 1 Samples S5, D1, D2, D3, D4, C1, C2 and C3 are expressed as mg/100 g “as is”. Samples S1, S2, S3, S4, S6, S7, S8 and S9 are expressed as mg/100 g reconstituted final product.

NOTE 2 SRM 1869, 1549a and 3233 are trade names of products supplied by NIST. This information is given for the convenience of users of this document and does not constitute an endorsement by ISO or IDF of the product named. Equivalent products may be used if they can be shown to lead to the same results.

NOTE 3 M-0142, MO-0614 and CA-0904 are products supplied by Muva Kempton GmbH. This information is given for the convenience of users of this document and does not constitute an endorsement by ISO or IDF of the product named. Equivalent products may be used if they can be shown to lead to the same results.

Table A.13 — Precision data for proline

Sample	S1 ^a	S2 ^b	S3 ^c	S4 ^d	S5 ^e	S6 ^f	S7 ^g	S8 ^h	S9 ⁱ	D1 ^j	D2 ^k	D3 ^l	D4 ^m	C1 ⁿ	C2 ^o	C3 ^p
Year of interlaboratory test	2020	2020	2020	2020	2020	2020	2020	2020	2020	2020	2020	2020	2020	2020	2020	2020
Number of laboratories retained after eliminating outliers	10	10	10	10	9	11	11	10	10	8	9	10	10	10	10	10
Number of outliers (laboratories)	1	1	1	1	0	0	0	1	1	1	1	0	0	0	0	0
Number of accepted results	11	11	11	11	9	11	11	11	11	9	10	10	10	10	10	10
Mean value, $\bar{x}$, mg/100 g	136,2	79,1	71,8	60,4	113,8	100,4	73,3	140,7	128,5	334,0	635,4	9 892,8	2 404,1	859,0	902,3	671,9
Repeatability standard deviation, s_r , mg/100 g	1,52	0,97	1,15	1,38	1,99	0,90	1,69	1,42	2,34	2,58	11,07	112,57	37,75	10,41	27,87	19,89
Reproducibility standard deviation, s_R , mg/100 g	4,45	2,85	1,50	2,27	5,23	3,12	2,39	4,14	3,92	12,74	22,07	466,92	87,63	28,02	29,87	35,65
Coefficient of variation of repeatability, $C_{V,r}$, %	1,12	1,22	1,61	2,29	1,75	0,90	2,31	1,01	1,82	0,77	1,74	1,14	1,57	1,21	3,09	2,96
Coefficient of variation of reproducibility, $C_{V,R}$, %	3,27	3,60	2,08	3,76	4,59	3,11	3,26	2,94	3,05	3,81	3,47	4,72	3,65	3,26	3,31	5,31
Repeatability limit, r [$r = 2,8 \times s_r$], mg/100 g	4,27	2,71	3,23	3,87	5,57	2,52	4,74	3,98	6,56	7,22	30,99	315,21	105,69	29,14	78,04	55,70
Reproducibility limit, R [$R = 2,8 \times s_R$], mg/100 g	12,47	7,98	4,19	6,36	14,63	8,73	6,69	11,58	10,99	35,67	61,81	1 307,39	245,37	78,45	83,65	99,82
Recovery, %	97															
HorRat value	0,61	0,61	0,35	0,62	0,83	0,55	0,55	0,55	0,56	0,81	0,81	1,67	1,04	0,80	0,82	1,25

^a SRM 1869 (infant/adult nutritional formula II); ^b IF, partially hydrolysed, milk-based; ^c IF, partially hydrolysed, soy-based; ^d IF, elemental, amino-acid-based; ^e IF, ready-to-feed, milk-based; ^f IF, milk-based; ^g IF, soy-based; ^h Toddler formula, milk-based; ⁱ Adult nutritional powder, low fat; ^j M-0142 (UHT skimmed milk); ^k MO-0614 (whey powder); ^l CA-0904 (sodium caseinate); ^m SRM 1549a (whole milk powder); ⁿ Bran pet food; ^o Dry pet food; ^p SRM 3233 (fortified breakfast cereal).

NOTE 1 Samples S5, D1, D2, D3, D4, C1, C2 and C3 are expressed as mg/100 g "as is". Samples S1, S2, S3, S4, S6, S7, S8 and S9 are expressed as mg/100 g reconstituted final product.

NOTE 2 SRM 1869, 1549a and 3233 are trade names of products supplied by NIST. This information is given for the convenience of users of this document and does not constitute an endorsement by ISO or IDF of the product named. Equivalent products may be used if they can be shown to lead to the same results.

NOTE 3 M-0142, MO-0614 and CA-0904 are products supplied by Muva Kempten GmbH. This information is given for the convenience of users of this document and does not constitute an endorsement by ISO or IDF of the product named. Equivalent products may be used if they can be shown to lead to the same results.

Table A.14 — Precision data for serine

Sample	S1 ^a	S2 ^b	S3 ^c	S4 ^d	S5 ^e	S6 ^f	S7 ^g	S8 ^h	S9 ⁱ	D1 ^j	D2 ^k	D3 ^l	D4 ^m	C1 ⁿ	C2 ^o	C3 ^p
Year of interlaboratory test	2020	2020	2020	2020	2020	2020	2020	2020	2020	2020	2020	2020	2020	2020	2020	2020
Number of laboratories retained after eliminating outliers	10	11	9	10	9	11	11	11	10	8	10	10	10	10	8	9
Number of outliers (laboratories)	1	0	2	1	0	0	0	0	1	1	0	0	0	0	2	1
Number of accepted results	11	11	11	11	9	11	11	11	11	9	10	10	10	10	10	10
Mean value, $\bar{x}$, mg/100 g	89,6	72,1	74,8	56,0	78,9	67,9	75,8	89,3	70,4	190,4	554,8	5 378,2	1 400,8	471,4	780,5	335,9
Repeatability standard deviation, s_r , mg/100 g	1,75	1,09	1,96	1,12	1,23	0,87	1,84	1,59	0,85	2,51	13,07	77,98	29,32	8,74	15,52	6,95
Reproducibility standard deviation, s_R , mg/100 g	5,75	4,10	4,86	3,55	4,13	3,84	4,64	4,85	3,76	12,85	50,83	443,58	87,78	37,87	26,94	18,72
Coefficient of variation of repeatability, $C_{V,r}$, %	1,96	1,51	2,62	1,99	1,56	1,28	2,43	1,78	1,20	1,32	2,36	1,45	2,09	1,85	1,99	2,07
Coefficient of variation of reproducibility, $C_{V,R}$, %	6,42	5,69	6,49	6,34	5,23	5,65	6,12	5,42	5,33	6,75	9,16	8,25	6,27	8,03	3,45	5,57
Repeatability limit, r [$r = 2,8 \times s_r$], mg/100 g	4,91	3,05	5,49	3,12	3,44	2,44	5,16	4,46	2,37	7,03	36,60	218,35	82,09	24,46	43,45	19,46
Reproducibility limit, R [$R = 2,8 \times s_R$], mg/100 g	16,09	11,49	13,61	9,94	11,56	10,75	12,98	13,57	10,52	35,98	142,32	1 242,03	245,78	106,04	75,43	52,40
Recovery, %	100												99			91
HorRat value	1,12	0,96	1,10	1,03	0,89	0,94	1,04	0,94	0,89	1,31	2,10	2,66	1,65	1,79	0,83	1,18

^a SRM 1869 (infant/adult nutritional formula II); ^b IF, partially hydrolysed, milk-based; ^c IF, partially hydrolysed, soy-based; ^d IF, elemental, amino-acid-based; ^e IF, ready-to-feed, milk-based; ^f IF, milk-based; ^g IF, soy-based; ^h Toddler formula, milk-based; ⁱ Adult nutritional powder, low fat; ^j M-0142 (UHT skimmed milk); ^k MO-0614 (whey powder); ^l CA-0904 (sodium caseinate); ^m SRM 1549a (whole milk powder); ⁿ Bran pet food; ^o Dry pet food; ^p SRM 3233 (fortified breakfast cereal).

NOTE 1 Samples S5, D1, D2, D3, D4, C1, C2 and C3 are expressed as mg/100 g “as is”. Samples S1, S2, S3, S4, S6, S7, S8 and S9 are expressed as mg/100 g reconstituted final product.

NOTE 2 SRM 1869, 1549a and 3233 are trade names of products supplied by NIST. This information is given for the convenience of users of this document and does not constitute an endorsement by ISO or IDF of the product named. Equivalent products may be used if they can be shown to lead to the same results.

NOTE 3 M-0142, MO-0614 and CA-0904 are products supplied by Muva Kempton GmbH. This information is given for the convenience of users of this document and does not constitute an endorsement by ISO or IDF of the product named. Equivalent products may be used if they can be shown to lead to the same results.

Table A.15 — Precision data for threonine

Sample	S1 ^a	S2 ^b	S3 ^c	S4 ^d	S5 ^e	S6 ^f	S7 ^g	S8 ^h	S9 ⁱ	D1 ^j	D2 ^k	D3 ^l	D4 ^m	C1 ⁿ	C2 ^o	C3 ^p
Year of interlaboratory test	2020	2020	2020	2020	2020	2020	2020	2020	2020	2020	2020	2020	2020	2020	2020	2020
Number of laboratories retained after eliminating outliers	9	10	10	9	9	11	10	11	10	8	9	10	10	10	9	10
Number of outliers (laboratories)	2	1	1	2	0	0	1	0	1	1	1	0	0	0	1	0
Number of accepted results	11	11	11	11	9	11	11	11	11	9	10	10	10	10	10	10
Mean value, $\bar{x}$, mg/100 g	76,9	95,1	53,7	77,7	86,1	67,5	55,1	81,8	68,9	150,9	719,3	4 048,8	1 099,2	319,9	611,6	217,2
Repeatability standard deviation, s_r , mg/100 g	0,54	0,82	0,79	2,00	1,48	0,62	0,73	1,41	0,70	1,54	12,35	55,32	21,25	4,93	11,64	5,07
Reproducibility standard deviation, s_R , mg/100 g	4,03	3,72	2,43	3,65	3,99	2,89	2,32	3,18	2,77	6,09	31,31	245,13	45,53	12,03	18,09	9,31
Coefficient of variation of repeatability, $C_{V,r}$, %	0,71	0,86	1,48	2,57	1,72	0,91	1,32	1,72	1,01	1,02	1,72	1,37	1,93	1,54	1,90	2,33
Coefficient of variation of reproducibility, $C_{V,R}$, %	5,24	3,91	4,52	4,69	4,63	4,28	4,22	3,89	4,03	4,04	4,35	6,05	4,14	3,76	2,96	4,29
Repeatability limit, r [$r = 2,8 \times s_r$], mg/100 g	1,52	2,29	2,22	5,59	4,15	1,73	2,03	3,94	1,95	4,32	34,59	154,91	59,51	13,79	32,58	14,18
Reproducibility limit, R [$R = 2,8 \times s_R$], mg/100 g	11,28	10,41	6,80	10,21	11,17	8,10	6,50	8,91	7,77	17,06	87,67	686,35	127,47	33,67	50,66	26,08
Recovery, %	99												101			92
HorRat value	0,89	0,69	0,73	0,80	0,80	0,71	0,68	0,67	0,67	0,76	1,04	1,87	1,05	0,79	0,69	0,85

^a SRM 1869 (infant/adult nutritional formula II); ^b IF, partially hydrolysed, milk-based; ^c IF, partially hydrolysed, soy-based; ^d IF, elemental, amino-acid-based; ^e IF, ready-to-feed, milk-based; ^f IF, milk-based; ^g IF, soy-based; ^h Toddler formula, milk-based; ⁱ Adult nutritional powder, low fat; ^j M-0142 (UHT skimmed milk); ^k MO-0614 (whey powder); ^l CA-0904 (sodium caseinate); ^m SRM 1549a (whole milk powder); ⁿ Bran pet food; ^o Dry pet food; ^p SRM 3233 (fortified breakfast cereal).

NOTE 1 Samples S5, D1, D2, D3, D4, C1, C2 and C3 are expressed as mg/100 g "as is". Samples S1, S2, S3, S4, S6, S7, S8 and S9 are expressed as mg/100 g reconstituted final product.

NOTE 2 SRM 1869, 1549a and 3233 are trade names of products supplied by NIST. This information is given for the convenience of users of this document and does not constitute an endorsement by ISO or IDF of the product named. Equivalent products may be used if they can be shown to lead to the same results.

NOTE 3 M-0142, MO-0614 and CA-0904 are products supplied by Muva Kempten GmbH. This information is given for the convenience of users of this document and does not constitute an endorsement by ISO or IDF of the product named. Equivalent products may be used if they can be shown to lead to the same results.

Table A.16 — Precision data for tyrosine

Sample	S1 ^a	S2 ^b	S3 ^c	S4 ^d	S5 ^e	S6 ^f	S7 ^g	S8 ^h	S9 ⁱ	D1 ^j	D2 ^k	D3 ^l	D4 ^m	C1 ⁿ	C2 ^o	C3 ^p
Year of interlaboratory test	2020	2020	2020	2020	2020	2020	2020	2020	2020	2020	2020	2020	2020	2020	2020	2020
Number of laboratories retained after eliminating outliers	8	9	9	9	7	9	10	10	9	7	9	9	9	9	8	7
Number of outliers (laboratories)	2	1	1	1	1	1	0	0	1	1	0	0	0	0	1	2
Number of accepted results	10	10	10	10	8	10	10	10	10	8	9	9	9	9	9	9
Mean value, $\bar{x}$, mg/100 g	72,8	38,0	51,7	92,3	55,4	50,1	53,0	73,1	66,9	172,2	289,4	5 481,0	1 229,6	272,8	478,8	192,2
Repeatability standard deviation, s_r , mg/100 g	0,63	0,43	1,85	2,94	2,10	0,88	2,33	1,56	1,90	1,43	21,81	195,15	32,86	11,67	10,51	5,19
Reproducibility standard deviation, s_R , mg/100 g	3,70	2,11	2,45	4,42	2,86	3,08	3,20	3,12	2,86	6,07	30,59	384,91	70,98	34,42	23,51	7,32
Coefficient of variation of repeatability, $C_{V,r}$, %	0,86	1,12	3,58	3,18	3,79	1,76	4,40	2,13	2,84	0,83	7,54	3,56	2,67	4,28	2,19	2,70
Coefficient of variation of reproducibility, $C_{V,R}$, %	5,09	5,55	4,74	4,79	5,17	6,15	6,04	4,27	4,27	3,52	10,57	7,02	5,77	12,62	4,91	3,81
Repeatability limit, r [$r = 2,8 \times s_r$], mg/100 g	1,76	1,19	5,18	8,22	5,88	2,47	6,53	4,36	5,32	4,00	61,07	546,42	92,01	32,69	29,42	14,54
Reproducibility limit, R [$R = 2,8 \times s_R$], mg/100 g	10,37	5,91	6,87	12,38	8,02	8,63	8,97	8,74	8,01	16,98	85,64	1 077,74	198,74	96,37	65,82	20,51
Recovery, %	107												110			85
HorRat value	0,86	0,85	0,76	0,84	0,84	0,98	0,97	0,72	0,71	0,68	2,19	2,27	1,49	2,59	1,10	0,74

^a SRM 1869 (infant/adult nutritional formula II); ^b IF, partially hydrolysed, milk-based; ^c IF, partially hydrolysed, soy-based; ^d IF, elemental, amino-acid-based; ^e IF, ready-to-feed, milk-based; ^f IF, milk-based; ^g IF, soy-based; ^h Toddler formula, milk-based; ⁱ Adult nutritional powder, low fat; ^j M-0142 (UHT skimmed milk); ^k MO-0614 (whey powder); ^l CA-0904 (sodium caseinate); ^m SRM 1549a (whole milk powder); ⁿ Bran pet food; ^o Dry pet food; ^p SRM 3233 (fortified breakfast cereal).

NOTE 1 Samples S5, D1, D2, D3, D4, C1, C2 and C3 are expressed as mg/100 g “as is”. Samples S1, S2, S3, S4, S6, S7, S8 and S9 are expressed as mg/100 g reconstituted final product.

NOTE 2 SRM 1869, 1549a and 3233 are trade names of products supplied by NIST. This information is given for the convenience of users of this document and does not constitute an endorsement by ISO or IDF of the product named. Equivalent products may be used if they can be shown to lead to the same results.

NOTE 3 M-0142, MO-0614 and CA-0904 are products supplied by Muva Kempton GmbH. This information is given for the convenience of users of this document and does not constitute an endorsement by ISO or IDF of the product named. Equivalent products may be used if they can be shown to lead to the same results.

Table A.17 — Precision data for valine

Sample	S1 ^a	S2 ^b	S3 ^c	S4 ^d	S5 ^e	S6 ^f	S7 ^g	S8 ^h	S9 ⁱ	D1 ^j	D2 ^k	D3 ^l	D4 ^m	C1 ⁿ	C2 ^o	C3 ^p
Year of interlaboratory test	2020	2020	2020	2020	2020	2020	2020	2020	2020	2020	2020	2020	2020	2020	2020	2020
Number of laboratories retained after eliminating outliers	10	9	10	10	9	11	11	11	10	8	10	10	10	10	10	10
Number of outliers (laboratories)	1	2	1	1	0	0	0	0	1	1	0	0	0	0	0	0
Number of accepted results	11	11	11	11	9	11	11	11	11	9	10	10	10	10	10	10
Mean value, $\bar{x}$, mg/100 g	95,9	77,7	66,9	139,2	87,2	75,1	69,3	99,2	99,9	212,4	617,1	6088,8	1555,5	468,0	769,7	325,7
Repeatability standard deviation, s_r , mg/100 g	0,66	0,50	1,66	1,20	1,24	0,80	2,06	1,68	0,92	2,11	14,59	52,65	20,49	8,23	22,93	7,49
Reproducibility standard deviation, s_R , mg/100 g	4,44	3,16	3,56	5,61	4,40	2,38	2,88	3,63	3,40	10,89	31,08	342,90	88,39	20,67	31,93	17,54
Coefficient of variation of repeatability, $C_{V,r}$, %	0,69	0,65	2,48	0,86	1,42	1,07	2,97	1,70	0,92	0,99	2,36	0,86	1,32	1,76	2,98	2,30
Coefficient of variation of reproducibility, $C_{V,R}$, %	4,63	4,08	5,32	4,03	5,05	3,17	4,15	3,66	3,41	5,13	5,04	5,63	5,68	4,42	4,15	5,38
Repeatability limit, r [$r = 2,8 \times s_r$], mg/100 g	1,86	1,41	4,64	3,36	3,46	2,24	5,76	4,71	2,56	5,91	40,86	147,41	57,37	23,03	64,20	20,97
Reproducibility limit, R [$R = 2,8 \times s_R$], mg/100 g	12,44	8,86	9,97	15,72	12,33	6,67	8,05	10,16	9,52	30,49	87,02	960,12	247,49	57,89	89,40	49,11
Recovery, %	100												116			97
HorRat value	0,81	0,69	0,89	0,75	0,87	0,54	0,69	0,65	0,60	1,02	1,17	1,85	1,52	0,99	1,00	1,14

^a SRM 1869 (infant/adult nutritional formula II); ^b IF, partially hydrolysed, milk-based; ^c IF, partially hydrolysed, soy-based; ^d IF, elemental, amino-acid-based; ^e IF, ready-to-feed, milk-based; ^f IF, milk-based; ^g IF, soy-based; ^h Toddler formula, milk-based; ⁱ Adult nutritional powder, low fat; ^j M-0142 (UHT skimmed milk); ^k MO-0614 (whey powder); ^l CA-0904 (sodium caseinate); ^m SRM 1549a (whole milk powder); ⁿ Bran pet food; ^o Dry pet food; ^p SRM 3233 (fortified breakfast cereal).

NOTE 1 Samples S5, D1, D2, D3, D4, C1, C2 and C3 are expressed as mg/100 g "as is". Samples S1, S2, S3, S4, S6, S7, S8 and S9 are expressed as mg/100 g reconstituted final product.

NOTE 2 SRM 1869, 1549a and 3233 are trade names of products supplied by NIST. This information is given for the convenience of users of this document and does not constitute an endorsement by ISO or IDF of the product named. Equivalent products may be used if they can be shown to lead to the same results.

NOTE 3 M-0142, MO-0614 and CA-0904 are products supplied by Muva Kempten GmbH. This information is given for the convenience of users of this document and does not constitute an endorsement by ISO or IDF of the product named. Equivalent products may be used if they can be shown to lead to the same results.

Annex B
(informative)

Chromatogram examples




Figure B.1 — Chromatogram of amino acid standards




Figure B.2 — Chromatogram of infant formula

Bibliography

- [1] BARKHOLT, V., JENSEN, A.L. Amino acid analysis: Determination of cysteine plus half-cystine in proteins after hydrochloric acid hydrolysis with a disulphide compound as additive. *Anal. Biochem.* 1989, 177 pp. 318–322
- [2] JAUDZEMS, G., FUERER, C. Determination of Total Amino Acids in Infant Formulas, Adult Nutritionals, Dairy, and Cereal Matrixes by UHPLC-UV: Interlaboratory Validation Study, Final Action 2018.06. *J. AOAC Int.* 2022, 105(6) pp.1625–1639. doi: 10.1093/jaoacint/qsac083
- [3] ISO 5725-2:2019, *Accuracy (trueness and precision) of measurement methods and results — Part 2: Basic method for the determination of repeatability and reproducibility of a standard measurement method*
- [4] AOAC International. Collaborative Study Guidelines. *J. AOAC Int.* 1995, 78 pp. 143A–161A
- [5] AOAC International. AOAC SMPR 2014.013. *J. AOAC Int.* 1995, 98, 1094, revised on, 2018-08-25
- [6] Codex Standard 72-1981, *Standard for infant formula and formulas for special medical purposes intended for infants*

